



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90036

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/12/2017	17:00
2	CAL2	1	Water			7.01	09/12/2017	17:01
3	CAL3	1	Water			10.03	09/12/2017	17:02
4	ICV	1	Water			7.00	09/12/2017	17:03
5	CCV1	1	Water			2.01	09/12/2017	17:15
6	I5210-03	1	Solid	20.02	20	8.80	09/12/2017	17:16
7	I5216-03	1	Solid	20.03	20	8.85	09/12/2017	17:18
8	I5216-06	1	Solid	20.04	20	8.82	09/12/2017	17:20
9	I5216-09	1	Solid	20.02	20	8.93	09/12/2017	17:21
10	I5218-03	1	Solid	20.03	20	8.53	09/12/2017	17:22
11	I5222-04	1	Solid	20.05	20	9.25	09/12/2017	17:23
12	I5222-08	1	Solid	20.03	20	8.54	09/12/2017	17:25
13	I5222-08DUP	1	Solid	20.02	20	8.55	09/12/2017	17:26
14	CCV2	1	Water			12.02	09/12/2017	17:29

WORKLIST(Hardcopy Internal Chain)

2990036

WorkList Name : CORROSIONITY I5222

WorkList ID : 102833

Date : 9/12/2017 4:54:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5210-03	Corrosivity	Cool 4 deg C	THEP04	D41	SASP2P-TP-BM-2(0-5)COMP	09/11/2017	9045C
	Solid	I5216-03	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-02-091217-A	09/12/2017	9045C
	Solid	I5216-06	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-02-091217-B	09/12/2017	9045C
	Solid	I5216-09	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-02-091217-C	09/12/2017	9045C
	Solid	I5218-03	Corrosivity	Cool 4 deg C	PSEG05	D52	SU-02-091217	09/12/2017	9045C
	Solid	I5222-04	Corrosivity	Cool 4 deg C	PSEG01	D52	SP-1	09/12/2017	9040C
	Solid	I5222-08	Corrosivity	Cool 4 deg C	PSEG01	D52	SP-2	09/12/2017	9040C

Date/Time 09-12-17 11:54 AM
 Received by: JO
 Relinquished by: JO

Date/Time 09-12-17 6:30 PM
 Received by: UP
 Relinquished by: JO