



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90055

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/13/2017	15:10
2	CAL2	1	Water			7.00	09/13/2017	15:11
3	CAL3	1	Water			10.02	09/13/2017	15:12
4	ICV	1	Water			7.00	09/13/2017	15:13
5	CCV1	1	Water			2.01	09/13/2017	15:16
6	I5231-03	1	Solid	20.03	20	9.77	09/13/2017	15:17
7	I5236-03	1	Solid	20.04	20	8.04	09/13/2017	15:19
8	I5236-03DUP	1	Solid	20.02	20	8.05	09/13/2017	15:20
9	CCV2	1	Water			12.02	09/13/2017	15:22

2B 900 55

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY i5236

WorkList ID : 102877

Date : 9/13/2017 4:09:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2017	Solid	I5231-03	Corrosivity	Cool 4 deg C	THEP04	E21	SASP2P-TP-116B(0-5)COMP	09/12/2017	9045C
	Solid	I5236-03	Corrosivity	Cool 4 deg C	PSEG05	E52	TR-04-091317	09/13/2017	9045C

Date/Time 09-13-17 4:10 PM
 Received by: JD
 Relinquished by: JD

Date/Time 09-13-17 6:20 PM
 Received by: ED
 Relinquished by: JD