

ISM02.4


LB 90065

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45714



A= 0.0000e+000

B= 2.1966e-004

C= 7.3613e-003

Rho= 0.9998095

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.00	0.000	-0.063	-0.063	-322	0.000	-322					-
0.2	0.200	0.208	0.008	913	0.0 %	913					4
2.5	2.500	2.618	0.118	11885	0.0 %	11885					5
5.0	5.000	4.992	-0.008	22694	0.0 %	22694					0
7.5	7.500	7.397	-0.103	33640	0.0 %	33640					-1
10.0	10.000	10.049	0.049	45714	0.0 %	45714					0

LB
9/14/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-322	-		0 ISM02.4	PPB	9/14/2017 11:32	Std
	0.2 S0.2	913	-		0.2 ISM02.4	PPB	9/14/2017 11:35	Std
	2.5 S2.5	11885	-		2.5 ISM02.4	PPB	9/14/2017 11:37	Std
	5 S5	22694	-		5 ISM02.4	PPB	9/14/2017 11:39	Std
	7.5 S7.5	33640	-		7.5 ISM02.4	PPB	9/14/2017 11:44	Std
	10 S10	45714	-		10 ISM02.4	PPB	9/14/2017 11:52	Std
ICV	ICV	18281	4.0229 -		ISM02.4	PPB	9/14/2017 11:57	SMPL
ICB	ICB	-290	-0.0563 -		ISM02.4	PPB	9/14/2017 11:59	SMPL
CCV025	CCV025	22554	4.9615 -		ISM02.4	PPB	9/14/2017 12:01	SMPL
CCB025	CCB025	-223	-0.0416 -		ISM02.4	PPB	9/14/2017 12:04	SMPL
PB102315BL	PBW012	-104	-0.0155 -		ISM02.4	PPB	9/14/2017 12:06	SMPL
I5192-06	MBDD18	-215	-0.0399 -		ISM02.4	PPB	9/14/2017 12:08	SMPL
PB102316BL	PBW012	-191	-0.0346 -		ISM02.4	PPB	9/14/2017 12:10	SMPL
I5193-06	MBDD17	-237	-0.0447 -		ISM02.4	PPB	9/14/2017 12:13	SMPL
CCV026	CCV026	22573	4.9656 -		ISM02.4	PPB	9/14/2017 12:15	SMPL
CCB026	CCB026	-297	-0.0579 -		ISM02.4	PPB	9/14/2017 12:17	SMPL