


7471A

LB90069
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40203



A= 0.0000e+000

B= 2.4813e-004

C= 3.8286e-002

Rho= 0.9999468

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-38	0.000	-38				
0.2	0.200	0.223	0.023	743	0.0 %	743				
2.5	2.500	2.467	-0.033	9787	0.0 %	9787				
5.0	5.000	4.931	-0.069	19719	0.0 %	19719				
7.5	7.500	7.536	0.036	30218	0.0 %	30218	0			
10.0	10.000	10.014	0.014	40203	0.0 %	40203				

LB90069 INSTRUMENT ID ; CV1

Method: 7471A

Operator: Admin

Date of Analysis: 14 Sep 2017 12:31:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-38	-	0.00007471A	PPB		14 Sep 2017 12:33:51	S	Std
0.2 - 1	S0.2	743	-	0.20007471A	PPB		14 Sep 2017 12:36:08	S	Std
2.5 - 1	S2.5	9787	-	2.50007471A	PPB		14 Sep 2017 12:38:26	S	Std
5.0 - 1	S5	19719	-	5.00007471A	PPB		14 Sep 2017 12:40:44	S	Std
7.5 - 1	S7.5	30218	-	7.50007471A	PPB		14 Sep 2017 12:43:01	S	Std
10.0 - 1	S10	40203	-	10.00007471A	PPB		14 Sep 2017 12:47:51	S	Std
ICV56 - 1	ICV56	15866	3.9752	-7471A	PPB		14 Sep 2017 12:50:29	U	SMPL
ICB56 - 1	ICB56	-175	-0.0051	-7471A	PPB		14 Sep 2017 12:52:45	U	SMPL
CCV27 - 1	CCV27	19844	4.9623	-7471A	PPB		14 Sep 2017 12:55:01	U	SMPL
CCB27 - 1	CCB27	-140	0.0035	-7471A	PPB		14 Sep 2017 12:57:17	U	SMPL
CRA - 1	CRA	762	0.2274	-7471A	PPB		14 Sep 2017 12:59:33	U	SMPL
HighStd - 1	HighStd	40624	10.1185	-7471A	PPB		14 Sep 2017 13:01:49	U	SMPL
ChkStd - 1	ChkStd	29047	7.2459	-7471A	PPB		14 Sep 2017 13:04:05	U	SMPL
PB102339BL - 1	PBS	-339	-0.0458	-7471A	PPB		14 Sep 2017 13:06:23	U	SMPL
PB102339BS - 1	LCSS	14800	3.7107	-7471A	PPB		14 Sep 2017 13:08:41	U	SMPL
I4811-02 - 1	EP-18	919	0.2663	-7471A	PPB		14 Sep 2017 13:13:16	U	SMPL
I4811-03 - 1	EP-23	242	0.0983	-7471A	PPB		14 Sep 2017 13:15:34	U	SMPL
I4811-04 - 1	E4-24	202	0.0884	-7471A	PPB		14 Sep 2017 13:17:52	U	SMPL
I4811-04DUP - 1	E4-24DUP	154	0.0765	-7471A	PPB		14 Sep 2017 13:20:08	U	SMPL
CCV28 - 1	CCV28	19739	4.9362	-7471A	PPB		14 Sep 2017 13:22:24	U	SMPL
CCB28 - 1	CCB28	-187	-0.0081	-7471A	PPB		14 Sep 2017 13:24:40	U	SMPL
I4811-05 - 1	I4811-04MS	15197	3.8092	-7471A	PPB		14 Sep 2017 13:26:56	U	SMPL
I4811-06 - 1	I4811-04MSD	15154	3.7985	-7471A	PPB		14 Sep 2017 13:29:13	U	SMPL
I4811-07 - 1	DUP	158	0.0775	-7471A	PPB		14 Sep 2017 13:31:29	U	SMPL
I4811-04LX5 - 1		-23	0.0326	-7471A	PPB		14 Sep 2017 13:33:45	U	SMPL
I4811-04A - 1		16491	4.1303	-7471A	PPB		14 Sep 2017 13:36:01	U	SMPL
I4811-01 - 1	EP-12	4751	1.2172	-7471A	PPB		14 Sep 2017 13:38:18	U	SMPL
CCV29 - 1	CCV29	19855	4.9650	-7471A	PPB		14 Sep 2017 13:45:11	U	SMPL
CCB29 - 1	CCB29	-171	-0.0041	-7471A	PPB		14 Sep 2017 13:47:28	U	SMPL