


7470A

LB 90075
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35323



A= 0.0000e+000

B= 2.8102e-004

C= 4.0209e-002

Rho= 0.9997556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.060	0.060	72	0.000	72				
0.2	0.200	0.196	-0.004	553	0.0 %	553				
2.5	2.500	2.471	-0.029	8650	0.0 %	8650				
5.0	5.000	4.872	-0.128	17194	0.0 %	17194				
7.5	7.500	7.634	0.134	27022	0.0 %	27022				
10.0	10.000	9.967	-0.033	35323	0.0 %	35323				

LB90075 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Sep 2017 14:56:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.0000	7470A	PPB	14 Sep 2017 15:05:56	S	Std
0.2 - 1	50.2	553	-	0.2000	7470A	PPB	14 Sep 2017 15:08:05	S	Std
2.5 - 1	52.5	8650	-	2.5000	7470A	PPB	14 Sep 2017 15:10:14	S	Std
5.0 - 1	55	17194	-	5.0000	7470A	PPB	14 Sep 2017 15:12:22	S	Std
7.5 - 1	57.5	27022	-	7.5000	7470A	PPB	14 Sep 2017 15:14:30	S	Std
10.0 - 1	510	35323	-	10.0000	7470A	PPB	14 Sep 2017 15:16:37	S	Std
ICV57 - 1	ICV57	13677	3.8838	-	7470A	PPB	14 Sep 2017 15:21:18	U	SMPL
ICB57 - 1	ICB57	-46	0.0273	-	7470A	PPB	14 Sep 2017 15:23:24	U	SMPL
CCV30 - 1	CCV30	17484	4.9536	-	7470A	PPB	14 Sep 2017 15:25:31	U	SMPL
CCB30 - 1	CCB30	-148	-0.0014	-	7470A	PPB	14 Sep 2017 15:27:37	U	SMPL
CRA - 1	CRA	467	0.1714	-	7470A	PPB	14 Sep 2017 15:29:44	U	SMPL
HighStd - 1	HighStd	35768	10.0918	-	7470A	PPB	14 Sep 2017 15:31:51	U	SMPL
ChkStd - 1	ChkStd	24865	7.0278	-	7470A	PPB	14 Sep 2017 15:33:58	U	SMPL
PB102318BL - 1	PBW	-214	-0.0199	-	7470A	PPB	14 Sep 2017 15:36:07	U	SMPL
PB102318BS - 1	LCSW	13145	3.7343	-	7470A	PPB	14 Sep 2017 15:38:15	U	SMPL
PB102318TB - 1	PB102291TB	-12	0.0368	-	7470A	PPB	14 Sep 2017 15:40:22	U	SMPL
I5210-03 - 1	SASP2P-TP-BM-2(0-5)COMP	39	0.0512	-	7470A	PPB	14 Sep 2017 15:42:31	U	SMPL
I5216-03 - 1	EO-02-091217-A	-37	0.0298	-	7470A	PPB	14 Sep 2017 15:44:39	U	SMPL
I5216-06 - 1	EO-02-091217-B	-5	0.0388	-	7470A	PPB	14 Sep 2017 15:46:48	U	SMPL
I5216-09 - 1	EO-02-091217-C	-71	0.0203	-	7470A	PPB	14 Sep 2017 15:48:55	U	SMPL
CCV31 - 1	CCV31	17338	4.9126	-	7470A	PPB	14 Sep 2017 15:51:02	U	SMPL
CCB31 - 1	CCB31	-137	0.0017	-	7470A	PPB	14 Sep 2017 15:53:09	U	SMPL
I5218-03 - 1	SU-02-091217	-33	0.0309	-	7470A	PPB	14 Sep 2017 15:55:16	U	SMPL
I5222-04 - 1	SP-1	-28	0.0323	-	7470A	PPB	14 Sep 2017 15:57:23	U	SMPL
I5222-08 - 1	SP-2	-1	0.0399	-	7470A	PPB	14 Sep 2017 15:59:30	U	SMPL
I5222-08DUP - 1	SP-2DUP	39	0.0512	-	7470A	PPB	14 Sep 2017 16:01:36	U	SMPL
I5222-08MS - 1	SP-2MS	13398	3.8054	-	7470A	PPB	14 Sep 2017 16:03:44	U	SMPL
I5222-08MSD - 1	SP-2MSD	13220	3.7553	-	7470A	PPB	14 Sep 2017 16:05:52	U	SMPL
I5226-02 - 1	266-323	-180	-0.0104	-	7470A	PPB	14 Sep 2017 16:08:00	U	SMPL
I5222-08LX5 - 1		-35	0.0304	-	7470A	PPB	14 Sep 2017 16:10:08	U	SMPL
I5222-08A - 1		14219	4.0361	-	7470A	PPB	14 Sep 2017 16:12:16	U	SMPL
CCV32 - 1	CCV32	17369	4.9213	-	7470A	PPB	14 Sep 2017 16:14:23	U	SMPL
CCB32 - 1	CCB32	-138	0.0014	-	7470A	PPB	14 Sep 2017 16:16:30	U	SMPL