



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90093

Temp.: 20.3 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/14/2017	17:10
2	CAL2	1	Water			7.00	09/14/2017	17:11
3	CAL3	1	Water			10.03	09/14/2017	17:12
4	ICV	1	Water			7.00	09/14/2017	17:13
5	CCV1	1	Water			2.01	09/14/2017	17:15
6	I5251-01	1	Water			11.63	09/14/2017	17:17
7	I5251-01DUP	1	Water			11.64	09/14/2017	17:19
8	CCV2	1	Water			12.02	09/14/2017	17:21

2390093

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5251 WorkList ID : 102916 Date : 9/14/2017 4:43:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2017	Water	I5251-01-D	pH	Cool 4 deg C	SEAC01	E41	GROUNDWATER	09/14/2017	SM 4500-PH B

Date/Time 09.14.17 4:43 PM
Received by: JR
Relinquished by: CR

Date/Time 09.14.17 6:20 PM
Received by: CR
Relinquished by: JR