


7470A

LB 90094
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35273



A= 0.0000e+000

B= 2.8532e-004

C= 1.2484e-002

Rho= 0.9997852

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	86	0.000	86				
0.2	0.200	0.168	-0.032	544	0.0 %	544				
2.5	2.500	2.505	0.005	8735	0.0 %	8735				
5.0	5.000	5.064	0.064	17706	0.0 %	17706				
7.5	7.500	7.350	-0.150	25715	0.0 %	25715				
10.0	10.000	10.077	0.077	35273	0.0 %	35273				

LB90094 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Sep 2017 10:08:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.0000	7470A	PPB	15 Sep 2017 10:31:51	S	Std
0.2 - 1	50.2	544	-	0.2000	7470A	PPB	15 Sep 2017 10:34:00	S	Std
2.5 - 1	52.5	8735	-	2.5000	7470A	PPB	15 Sep 2017 10:36:08	S	Std
5.0 - 1	55	17706	-	5.0000	7470A	PPB	15 Sep 2017 10:38:16	S	Std
7.5 - 1	57.5	25715	-	7.5000	7470A	PPB	15 Sep 2017 10:40:24	S	Std
10.0 - 1	510	35273	-	10.0000	7470A	PPB	15 Sep 2017 10:42:32	S	Std
ICV59 - 1	ICV59	13982	4.0018	-	7470A	PPB	15 Sep 2017 10:47:54	U	SMPL
ICB59 - 1	ICB59	-50	-0.0018	-	7470A	PPB	15 Sep 2017 10:50:00	U	SMPL
CCV37 - 1	CCV37	17922	5.1260	-	7470A	PPB	15 Sep 2017 10:52:07	U	SMPL
CCB37 - 1	CCB37	-109	-0.0186	-	7470A	PPB	15 Sep 2017 10:54:13	U	SMPL
CRA - 1	CRA	487	0.1514	-	7470A	PPB	15 Sep 2017 10:56:20	U	SMPL
HighStd - 1	HighStd	35668	10.1893	-	7470A	PPB	15 Sep 2017 10:58:26	U	SMPL
ChkStd - 1	ChkStd	25252	7.2174	-	7470A	PPB	15 Sep 2017 11:00:33	U	SMPL
PB102346BL - 1	PBW	-226	-0.0520	-	7470A	PPB	15 Sep 2017 11:02:42	U	SMPL
PB102346BS - 1	LCSW	12778	3.6583	-	7470A	PPB	15 Sep 2017 11:04:51	U	SMPL
PB102346TB - 1	PB102322TB	-132	-0.0252	-	7470A	PPB	15 Sep 2017 11:06:59	U	SMPL
I5231-03 - 1	SASP2P-TP-116B(0-5)COMP	-57	-0.0038	-	7470A	PPB	15 Sep 2017 11:09:07	U	SMPL
I5236-03 - 1	TR-04-091317	-49	-0.0015	-	7470A	PPB	15 Sep 2017 11:11:15	U	SMPL
I5236-03DUP - 1	TR-04-091317DUP	-35	0.0025	-	7470A	PPB	15 Sep 2017 11:13:24	U	SMPL
I5236-03MS - 1	TR-04-091317MS	13336	3.8175	-	7470A	PPB	15 Sep 2017 11:15:30	U	SMPL
CCV38 - 1	CCV38	18039	5.1594	-	7470A	PPB	15 Sep 2017 11:17:37	U	SMPL
CCB38 - 1	CCB38	-71	-0.0078	-	7470A	PPB	15 Sep 2017 11:19:43	U	SMPL
I5236-03MSD - 1	TR-04-091317MSD	13367	3.8264	-	7470A	PPB	15 Sep 2017 11:21:51	U	SMPL
I5236-03LX5 - 1		-128	-0.0240	-	7470A	PPB	15 Sep 2017 11:23:58	U	SMPL
I5236-03A - 1		14089	4.0324	-	7470A	PPB	15 Sep 2017 11:26:04	U	SMPL
CCV39 - 1	CCV39	18001	5.1486	-	7470A	PPB	15 Sep 2017 11:28:11	U	SMPL
CCB39 - 1	CCB39	-134	-0.0257	-	7470A	PPB	15 Sep 2017 11:30:19	U	SMPL