


ISM02.4

LB90098
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41745



A= 0.0000e+000

B= 2.4200e-004

C= -1.5691e-003

Rho= 0.9998103

Accept=Accepted

slope
+ intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	0.027	0.027	116	0.000	116					-
0.2	0.200	0.193	-0.007	802	0.0 %	802					-4
2.5	2.500	2.551	0.051	10548	0.0 %	10548					2
5.0	5.000	4.953	-0.047	20475	0.0 %	20475					-1
7.5	7.500	7.376	-0.124	30487	0.0 %	30487					-2
10.0	10.000	10.101	0.101	41745	0.0 %	41745					1

203
9/15/12

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	116	-		0 ISM02.4	PPB	9/15/2017 14:03	Std
	0.2 S0.2	802	-		0.2 ISM02.4	PPB	9/15/2017 14:05	Std
	2.5 S2.5	10548	-		2.5 ISM02.4	PPB	9/15/2017 14:07	Std
	5 S5	20475	-		5 ISM02.4	PPB	9/15/2017 14:09	Std
	7.5 S7.5	30487	-		7.5 ISM02.4	PPB	9/15/2017 14:12	Std
	10 S10	41745	-		10 ISM02.4	PPB	9/15/2017 14:14	Std
ICV	ICV	15941	3.8561 -		ISM02.4	PPB	9/15/2017 14:17	SMPL
ICB	ICB	-103	-0.0265 -		ISM02.4	PPB	9/15/2017 14:19	SMPL
CCV040	CCV040	20329	4.918 -		ISM02.4	PPB	9/15/2017 14:21	SMPL
CCB040	CCB040	-100	-0.0258 -		ISM02.4	PPB	9/15/2017 14:24	SMPL
PB102347BL	PBW003	-14	-0.005 -		ISM02.4	PPB	9/15/2017 14:26	SMPL
I5241-01	MC0AA1	-218	-0.0543 -		ISM02.4	PPB	9/15/2017 14:28	SMPL
I5241-02	MC0AA1D	-210	-0.0524 -		ISM02.4	PPB	9/15/2017 14:31	SMPL
I5241-03	MC0AA1S	3898	0.9417 -		ISM02.4	PPB	9/15/2017 14:33	SMPL
CCV041	CCV041	20312	4.9138 -		ISM02.4	PPB	9/15/2017 14:35	SMPL
CCB041	CCB041	-121	-0.0309 -		ISM02.4	PPB	9/15/2017 14:37	SMPL