


7471A

LB90103
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42336



A= 0.0000e+000

B= 2.3830e-004

C= 6.3026e-002

Rho= 0.9995893

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.049	0.049	-58	0.000	-58				
0.2	0.200	0.204	0.004	592	0.0 %	592				
2.5	2.500	2.522	0.022	10320	0.0 %	10320				
5.0	5.000	4.974	-0.026	20610	0.0 %	20610				
7.5	7.500	7.299	-0.201	30364	0.0 %	30364	0			
10.0	10.000	10.152	0.152	42336	0.0 %	42336				

LB90103 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Sep 2017 15:28:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-58	-	0.0000	7471A	PPB	15 Sep 2017 15:29:36	S	Std
0.2 - 1	50.2	592	-	0.2000	7471A	PPB	15 Sep 2017 15:31:54	S	Std
2.5 - 1	52.5	10320	-	2.5000	7471A	PPB	15 Sep 2017 15:34:11	S	Std
5.0 - 1	35	20610	-	5.0000	7471A	PPB	15 Sep 2017 15:36:29	S	Std
7.5 - 1	57.5	30364	-	7.5000	7471A	PPB	15 Sep 2017 15:38:46	S	Std
10.0 - 1	510	42336	-	10.0000	7471A	PPB	15 Sep 2017 15:43:39	S	Std
ICV60 - 1	ICV60	15907	3.8536	-	7471A	PPB	15 Sep 2017 15:46:25	U	SMPL
ICB60 - 1	ICB60	-252	0.0030	-	7471A	PPB	15 Sep 2017 15:48:40	U	SMPL
CCV42 - 1	CCV42	20783	5.0155	-	7471A	PPB	15 Sep 2017 15:50:56	U	SMPL
CCB42 - 1	CCB42	-235	0.0070	-	7471A	PPB	15 Sep 2017 15:53:12	U	SMPL
CRA - 1	CRA	515	0.1857	-	7471A	PPB	15 Sep 2017 15:55:27	U	SMPL
HighStd - 1	HighStd	42604	10.2154	-	7471A	PPB	15 Sep 2017 15:57:43	U	SMPL
ChkStd - 1	ChkStd	28299	6.8066	-	7471A	PPB	15 Sep 2017 15:59:59	U	SMPL
PB102352BL - 1	PBS	-218	0.0111	-	7471A	PPB	15 Sep 2017 16:02:17	U	SMPL
PB102352BS - 1	LCSS	15504	3.7576	-	7471A	PPB	15 Sep 2017 16:04:35	U	SMPL
I5210-01 - 1	SASP2P-TP-BM-2(0-5)COMP	2487	0.6557	-	7471A	PPB	15 Sep 2017 16:06:52	U	SMPL
I5217-02 - 1	LP-SLUDGE	44536	10.6758	-	7471A	PPB	15 Sep 2017 16:09:09	U	SMPL
I5226-01 - 1	266-323	-130	0.0320	-	7471A	PPB	15 Sep 2017 16:11:26	U	SMPL
I5229-06 - 1	UT5400S16	3887	0.9893	-	7471A	PPB	15 Sep 2017 16:13:46	U	SMPL
I5229-07 - 1	UT5400S17	4024	1.0219	-	7471A	PPB	15 Sep 2017 16:16:02	U	SMPL
CCV43 - 1	CCV43	20517	4.9521	-	7471A	PPB	15 Sep 2017 16:18:18	U	SMPL
CCB43 - 1	CCB43	-230	0.0082	-	7471A	PPB	15 Sep 2017 16:20:34	U	SMPL
I5229-08 - 1	UT5400S18	4638	1.1682	-	7471A	PPB	15 Sep 2017 16:22:51	U	SMPL
I5229-09 - 1	UT5400S19	4340	1.0972	-	7471A	PPB	15 Sep 2017 16:25:07	U	SMPL
I5229-09DUP - 1	UT5400S19DUP	4201	1.0641	-	7471A	PPB	15 Sep 2017 16:27:22	U	SMPL
I5229-09MS - 1	UT5400S19MS	19040	4.6002	-	7471A	PPB	15 Sep 2017 16:29:39	U	SMPL
I5229-09MSD - 1	UT5400S19MSD	20388	4.9214	-	7471A	PPB	15 Sep 2017 16:31:55	U	SMPL
I5229-10 - 1	UT5400S20	3816	0.9724	-	7471A	PPB	15 Sep 2017 16:34:12	U	SMPL
I5229-11 - 1	UT5400S21	4952	1.2431	-	7471A	PPB	15 Sep 2017 16:36:29	U	SMPL
I5231-01 - 1	SASP2P-TP-116B(0-5)COMP	1513	0.4236	-	7471A	PPB	15 Sep 2017 16:38:47	U	SMPL
I5236-01 - 1	TR-04-091317	1789	0.4893	-	7471A	PPB	15 Sep 2017 16:41:05	U	SMPL
I5244-01 - 1	SASP2P-TP-106-X(0-5)COMP	2450	0.6469	-	7471A	PPB	15 Sep 2017 16:43:21	U	SMPL
CCV44 - 1	CCV44	20700	4.9957	-	7471A	PPB	15 Sep 2017 16:45:37	U	SMPL
CCB44 - 1	CCB44	-254	0.0025	-	7471A	PPB	15 Sep 2017 16:47:53	U	SMPL
I5249-01 - 1	ROLL-OFF1	45606	10.9307	-	7471A	PPB	15 Sep 2017 16:50:10	U	SMPL
I5250-01 - 1	OK-01-091417	1575	0.4383	-	7471A	PPB	15 Sep 2017 16:52:26	U	SMPL
I5252-01 - 1	TP-1-COMP	1227	0.3554	-	7471A	PPB	15 Sep 2017 16:54:45	U	SMPL
I5229-09LX5 - 1		678	0.2246	-	7471A	PPB	15 Sep 2017 16:57:02	U	SMPL
I5229-09A - 1		21487	5.1833	-	7471A	PPB	15 Sep 2017 16:59:19	U	SMPL
I5217-02DLX2 - 1		22729	5.4793	-	7471A	PPB	15 Sep 2017 17:01:36	U	SMPL
I5249-01DLX2 - 1		22852	5.5086	-	7471A	PPB	15 Sep 2017 17:03:54	U	SMPL
CCV45 - 1	CCV45	20532	4.9557	-	7471A	PPB	15 Sep 2017 17:06:11	U	SMPL
CCB45 - 1	CCB45	-108	0.0373	-	7471A	PPB	15 Sep 2017 17:12:51	U	SMPL