



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90122

Temp.: NA °C

SLOPE.: NA

Calibration Standards	Chemtech Log#
0.0-3.0 pH paper refill	W1931
3.0-5.5 pH paper refill	W1934
6.0-8.0 pH paper refill	W2106
8.2-9.8 pH paper refill	W1938
Test Paper, PH Short Range 9.0/10.0	W1939
10.0-12.0 pH paper refill	W1940
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER, COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.0	09/18/2017	09:30
2	CAL2	1	Water			7.0	09/18/2017	09:31
3	CAL3	1	Water			10.0	09/18/2017	09:33
4	ICV	1	Water			7.0	09/18/2017	09:35
5	CCV1	1	Water			2.0	09/18/2017	09:36
6	I5274-22	1	Water			8.2	09/18/2017	09:37
7	CCV2	1	Water			12.0	09/18/2017	09:38