



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90127

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/18/2017	10:50
2	CAL2	1	Water			7.01	09/18/2017	10:51
3	CAL3	1	Water			10.04	09/18/2017	10:53
4	ICV	1	Water			7.00	09/18/2017	11:00
5	CCV1	1	Water			2.01	09/18/2017	11:02
6	I5219-04	1	Solid	20.2	20	6.74	09/18/2017	11:04
7	I5219-04DUP	1	Solid	20.4	20	6.75	09/18/2017	11:05
8	CCV2	1	Water			12.01	09/18/2017	11:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5119

WorkList ID : 102987

Date : 9/18/2017 10:38:08 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5219-04-13	pH	Cool 4 deg C	TULL02	E21	FK-PS-01-015	09/12/2017	9045C

Date/Time 09.18.17 10:40 AM
 Received by: SD
 Relinquished by: SD

Date/Time 09.18.17 11:45 AM
 Received by: CR
 Relinquished by: SD