


7470A

LB 90129
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35198



A= 0.0000e+000

B= 2.8543e-004

C= -2.0017e-002

Rho= 0.9999500

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-52	0.000	-52				
0.2	0.200	0.209	0.009	801	0.0 %	801				
2.5	2.500	2.560	0.060	9040	0.0 %	9040				
5.0	5.000	4.989	-0.011	17548	0.0 %	17548				
7.5	7.500	7.451	-0.049	26174	0.0 %	26174				
10.0	10.000	10.027	0.027	35198	0.0 %	35198				

LB90129 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Sep 2017 13:06:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-52	-	0.0000	7470A	PPB	18 Sep 2017 13:52:22	S	Std
0.2 - 1	50.2	801	-	0.2000	7470A	PPB	18 Sep 2017 13:54:30	S	Std
2.5 - 1	50.5	9040	-	2.5000	7470A	PPB	18 Sep 2017 13:56:39	S	Std
5.0 - 1	55	17548	-	5.0000	7470A	PPB	18 Sep 2017 13:58:47	S	Std
7.5 - 1	50.5	26174	-	7.5000	7470A	PPB	18 Sep 2017 14:00:55	S	Std
10.0 - 1	510	35198	-	10.0000	7470A	PPB	18 Sep 2017 14:05:40	S	Std
ICV62 - 1	ICV62	13708	3.8926	-7470A	PPB	18 Sep 2017 14:08:41	U	SMPL	
ICB62 - 1	ICB62	-229	-0.0854	-7470A	PPB	18 Sep 2017 14:10:47	U	SMPL	
CCV50 - 1	CCV50	17672	5.0241	-7470A	PPB	18 Sep 2017 14:12:53	U	SMPL	
CCB50 - 1	CCB50	-188	-0.0737	-7470A	PPB	18 Sep 2017 14:15:00	U	SMPL	
CRA - 1	CRA	836	0.2186	-7470A	PPB	18 Sep 2017 14:17:06	U	SMPL	
HighStd - 1	HighStd	35349	10.0696	-7470A	PPB	18 Sep 2017 14:19:13	U	SMPL	
ChkStd - 1	ChkStd	24829	7.0669	-7470A	PPB	18 Sep 2017 14:21:20	U	SMPL	
PB102365BL - 1	PBW	-184	-0.0725	-7470A	PPB	18 Sep 2017 14:25:54	U	SMPL	
PB102365BS - 1	LCSW	14205	4.0345	-7470A	PPB	18 Sep 2017 14:28:01	U	SMPL	
PB102365TB - 1	PB102350TB	-180	-0.0714	-7470A	PPB	18 Sep 2017 14:30:08	U	SMPL	
I5249-02 - 1	ROLL-OFF1	-40	-0.0314	-7470A	PPB	18 Sep 2017 14:32:17	U	SMPL	
I5250-03 - 1	OK-01-091417	-69	-0.0397	-7470A	PPB	18 Sep 2017 14:34:25	U	SMPL	
I5252-02 - 1	TP-1-COMP	-116	-0.0531	-7470A	PPB	18 Sep 2017 14:36:33	U	SMPL	
I5252-02DUP - 1	TP-1-COMPDUP	-88	-0.0451	-7470A	PPB	18 Sep 2017 14:48:36	U	SMPL	
CCV51 - 1	CCV51	17409	4.9490	-7470A	PPB	18 Sep 2017 14:50:42	U	SMPL	
CCB51 - 1	CCB51	-261	-0.0945	-7470A	PPB	18 Sep 2017 14:52:49	U	SMPL	
I5252-02MS - 1	TP-1-COMPMS	13451	3.8193	-7470A	PPB	18 Sep 2017 14:54:55	U	SMPL	
I5252-02MSD - 1	TP-1-COMPMSD	13255	3.7633	-7470A	PPB	18 Sep 2017 14:57:02	U	SMPL	
PB102368BL - 1	PBW	-161	-0.0660	-7470A	PPB	18 Sep 2017 14:59:08	U	SMPL	
PB102368BS - 1	LCSW	13409	3.8073	-7470A	PPB	18 Sep 2017 15:01:15	U	SMPL	
I5259-01 - 1	20170603-COMP-D	-180	-0.0714	-7470A	PPB	18 Sep 2017 15:06:45	U	SMPL	
I5259-02 - 1	20170602-COMP-C	-218	-0.0822	-7470A	PPB	18 Sep 2017 15:08:53	U	SMPL	
I5259-03 - 1	20170601-COMP-B	-56	-0.0360	-7470A	PPB	18 Sep 2017 15:11:01	U	SMPL	
I5259-04 - 1	20170600-COMP-A	-184	-0.0725	-7470A	PPB	18 Sep 2017 15:13:10	U	SMPL	
I5259-05 - 1	20170603-COMP-D	-162	-0.0663	-7470A	PPB	18 Sep 2017 15:15:16	U	SMPL	
CCV52 - 1	CCV52	17646	5.0167	-7470A	PPB	18 Sep 2017 15:17:23	U	SMPL	
CCB52 - 1	CCB52	-206	-0.0788	-7470A	PPB	18 Sep 2017 15:19:30	U	SMPL	
I5259-06 - 1	20170602-COMP-C	-195	-0.0757	-7470A	PPB	18 Sep 2017 15:21:37	U	SMPL	
I5259-07 - 1	20170601-COMP-B	-62	-0.0377	-7470A	PPB	18 Sep 2017 15:23:44	U	SMPL	
I5259-08 - 1	20170600-COMP-A	-178	-0.0708	-7470A	PPB	18 Sep 2017 15:25:51	U	SMPL	
I5267-01 - 1	20170633-FROM-WIER	-196	-0.0760	-7470A	PPB	18 Sep 2017 15:27:58	U	SMPL	
I5267-01DUP - 1	20170633-FROM-WIERDUP	-176	-0.0703	-7470A	PPB	18 Sep 2017 15:30:06	U	SMPL	
I5267-01MS - 1	20170633-FROM-WIERMS	15793	4.4878	-7470A	PPB	18 Sep 2017 15:32:13	U	SMPL	
I5267-01MSD - 1	20170633-FROM-WIERMSD	15470	4.3956	-7470A	PPB	18 Sep 2017 15:34:21	U	SMPL	
I5267-01LX5 - 1		-227	-0.0848	-7470A	PPB	18 Sep 2017 15:36:29	U	SMPL	
I5267-01A - 1		16344	4.6450	-7470A	PPB	18 Sep 2017 15:38:38	U	SMPL	
I5217-01DLX5 - 1	LP-WATER	24611	7.0047	-7470A	PPB	18 Sep 2017 15:40:45	U	SMPL	
CCV53 - 1	CCV53	17399	4.9462	-7470A	PPB	18 Sep 2017 15:42:52	U	SMPL	
CCB53 - 1	CCB53	-280	-0.0999	-7470A	PPB	18 Sep 2017 15:45:00	U	SMPL	
I5252-02LX5 - 1		-128	-0.0566	-7470A	PPB	18 Sep 2017 15:47:08	U	SMPL	
I5252-02A - 1		14735	4.1858	-7470A	PPB	18 Sep 2017 15:49:15	U	SMPL	
CCV54 - 1	CCV54	17516	4.9796	-7470A	PPB	18 Sep 2017 15:51:22	U	SMPL	
CCB54 - 1	CCB54	-259	-0.0939	-7470A	PPB	18 Sep 2017 15:53:30	U	SMPL	