



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90137

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/18/2017	16:10
2	CAL2	1	Water			7.01	09/18/2017	16:11
3	CAL3	1	Water			10.03	09/18/2017	16:13
4	ICV	1	Water			7.00	09/18/2017	16:16
5	CCV1	1	Water			2.01	09/18/2017	16:17
6	I5280-01	1	Solid	20.02	20	6.38	09/18/2017	16:19
7	I5280-01DUP	1	Solid	20.04	20	6.39	09/18/2017	16:20
8	I5280-02	1	Solid	20.01	20	7.38	09/18/2017	16:21
9	CCV2	1	Water			12.02	09/18/2017	16:22

239013+

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i5280

WorkList ID : 102995

Date : 9/18/2017 3:59:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5280-01 - E	pH	Cool 4 deg C	LABE01	H12	BIOCELL-1	09/13/2017	9045C
	Solid	I5280-02 - E	pH	Cool 4 deg C	LABE01	H12	BIOCELL-2	09/13/2017	9045C

Date/Time 09-18-17 4:00 PM
 Received by: JP
 Relinquished by: J

Date/Time 09-18-17 5:00 PM
 Received by: Q
 Relinquished by: Si