



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90144

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.04	09/18/2017	18:00
2	CAL2	1	Water			7.01	09/18/2017	18:01
3	CAL3	1	Water			10.03	09/18/2017	18:03
4	ICV	1	Water			7.00	09/18/2017	18:05
5	CCV1	1	Water			2.02	09/18/2017	18:07
6	I5285-06	1	Solid	20.02	20	8.80	09/18/2017	18:09
7	I5285-06DUP	1	Solid	20.03	20	8.81	09/18/2017	18:11
8	I5285-07	1	Solid	20.04	20	9.05	09/18/2017	18:13
9	I5285-08	1	Solid	20.01	20	8.77	09/18/2017	18:15
10	I5285-09	1	Solid	20.03	20	8.96	09/18/2017	18:16
11	I5285-10	1	Solid	20.02	20	9.04	09/18/2017	18:19
12	I5288-02	1	Solid	20.03	20	8.94	09/18/2017	18:22
13	I5291-02	1	Solid	20.01	20	9.37	09/18/2017	18:23
14	I5292-03	1	Solid	20.02	20	6.35	09/18/2017	18:25
15	CCV2	1	Water			12.01	09/18/2017	18:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5285

WorkList ID : 102991

Date : 9/18/2017 1:50:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5285-06	Corrosivity	Cool 4 deg C	AREC01	H21	CSTP-1	09/18/2017	9045C
	Solid	I5285-07	Corrosivity	Cool 4 deg C	AREC01	H21	CSTP-2	09/18/2017	9045C
	Solid	I5285-08	Corrosivity	Cool 4 deg C	AREC01	H21	CSTP-3	09/18/2017	9045C
	Solid	I5285-09	Corrosivity	Cool 4 deg C	AREC01	H21	CSTP-4	09/18/2017	9045C
	Solid	I5285-10	Corrosivity	Cool 4 deg C	AREC01	H21	CSTP-5	09/18/2017	9045C
09/25/2017	Solid	I5288-02	Corrosivity	Cool 4 deg C	PSEG02	G52	DRUM-A-OILYSTONES	09/18/2017	9045C
09/25/2017	Solid	I5291-02	Corrosivity	Cool 4 deg C	PSEG02	G62	OILY-STONE	09/18/2017	9045C
09/21/2017	Solid	I5292-03	Corrosivity	Cool 4 deg C	PSEG05	G62	BU-01-091817	09/18/2017	9045C

Date/Time 09-18-17 11:00 AM
 Received by: JV
 Relinquished by: CR

Date/Time 09-18-17 5:30 PM
 Received by: CP
 Relinquished by: JH