


7471A

LB 90149
INSTRUMENT ID: CY1

Linear

 μ Abs.:

42364



A= 0.0000e+000

B= 2.3561e-004

C= -1.8735e-002

Rho= 0.9999737

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-17	0.000	-17				
0.2	0.200	0.212	0.012	978	0.0 %	978				
2.5	2.500	2.497	-0.003	10676	0.0 %	10676				
5.0	5.000	5.005	0.005	21323	0.0 %	21323				
7.5	7.500	7.547	0.047	32111	0.0 %	32111	0			
10.0	10.000	9.963	-0.037	42364	0.0 %	42364				

LB90149 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 19 Sep 2017 11:52:57					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-17	-	0.0000	7471A	PPB	19 Sep 2017 11:57:09	S	Std
0.2 - 1	50.2	978	-	0.2000	7471A	PPB	19 Sep 2017 11:59:27	S	Std
2.5 - 1	52.5	10676	-	2.5000	7471A	PPB	19 Sep 2017 12:01:44	S	Std
5.0 - 1	55	21323	-	5.0000	7471A	PPB	19 Sep 2017 12:04:02	S	Std
7.5 - 1	57.5	32111	-	7.5000	7471A	PPB	19 Sep 2017 12:06:19	S	Std
10.0 - 1	59	42364	-	10.0000	7471A	PPB	19 Sep 2017 12:08:36	S	Std
ICV63 - 1	ICV63	17012	3.9894	-	7471A	PPB	19 Sep 2017 12:11:18	U	SMPL
ICB63 - 1	ICB63	-239	-0.0750	-	7471A	PPB	19 Sep 2017 12:13:35	U	SMPL
CCV55 - 1	CCV55	20124	4.7226	-	7471A	PPB	19 Sep 2017 12:15:51	U	SMPL
CCB55 - 1	CCB55	-172	-0.0593	-	7471A	PPB	19 Sep 2017 12:18:07	U	SMPL
CRA - 1	CRA	959	0.2072	-	7471A	PPB	19 Sep 2017 12:20:23	U	SMPL
HighStd - 1	HighStd	42947	10.0999	-	7471A	PPB	19 Sep 2017 12:22:39	U	SMPL
ChkStd - 1	ChkStd	30173	7.0902	-	7471A	PPB	19 Sep 2017 12:24:55	U	SMPL
PB102427BL - 1	PBS	-324	-0.0951	-	7471A	PPB	19 Sep 2017 12:27:13	U	SMPL
PB102427BS - 1	LCSS	15671	3.6735	-	7471A	PPB	19 Sep 2017 12:29:31	U	SMPL
I5253-01 - 1	FK-BPM-01-C7-C8-D7-D8	1159	0.2543	-	7471A	PPB	19 Sep 2017 12:31:48	U	SMPL
I5254-01 - 1	FK-BPM-01-E5-E6-F5-F6	1370	0.3040	-	7471A	PPB	19 Sep 2017 12:34:05	U	SMPL
I5276-01 - 1	SASP2P-ENV-117-4(0-5)	1468	0.3271	-	7471A	PPB	19 Sep 2017 12:36:23	U	SMPL
I5276-02 - 1	SASP2P-ENV-117-4(5-10)	91	0.0027	-	7471A	PPB	19 Sep 2017 12:38:41	U	SMPL
I5276-03 - 1	SASP2P-ENV-125-14(0-5)	9798	2.2897	-	7471A	PPB	19 Sep 2017 12:40:57	U	SMPL
CCV56 - 1	CCV56	20322	4.7693	-	7471A	PPB	19 Sep 2017 12:43:12	U	SMPL
CCB56 - 1	CCB56	-194	-0.0644	-	7471A	PPB	19 Sep 2017 12:45:29	U	SMPL
I5276-04 - 1	SASP2P-ENV-125-14(5-10)	1680	0.3771	-	7471A	PPB	19 Sep 2017 12:47:46	U	SMPL
I5276-05 - 1	SASP2P-ENV-125-3(0-5)	10511	2.4577	-	7471A	PPB	19 Sep 2017 12:50:02	U	SMPL
I5276-06 - 1	SASP2P-ENV-118-3(0-5)	880	0.1886	-	7471A	PPB	19 Sep 2017 12:52:18	U	SMPL
I5276-07 - 1	SASP2P-ENV-118-3(5-10)	208	0.0303	-	7471A	PPB	19 Sep 2017 12:54:34	U	SMPL
I5276-07DUP - 1	SASP2P-ENV-118-3(5-10)DUP	217	0.0324	-	7471A	PPB	19 Sep 2017 12:56:51	U	SMPL
I5276-07MS - 1	SASP2P-ENV-118-3(5-10)MS	16149	3.7861	-	7471A	PPB	19 Sep 2017 12:59:08	U	SMPL
I5276-07MSD - 1	SASP2P-ENV-118-3(5-10)MSD	16048	3.7623	-	7471A	PPB	19 Sep 2017 13:01:25	U	SMPL
I5285-01 - 1	CSTP-1	54148	12.7389	-	7471A	PPB	19 Sep 2017 13:03:44	U	SMPL
I5285-02 - 1	CSTP-2	35840	8.4254	-	7471A	PPB	19 Sep 2017 13:06:02	U	SMPL
I5285-03 - 1	CSTP-3	28994	6.8124	-	7471A	PPB	19 Sep 2017 13:08:19	U	SMPL
CCV57 - 1	CCV57	20245	4.7511	-	7471A	PPB	19 Sep 2017 13:10:37	U	SMPL
CCB57 - 1	CCB57	-232	-0.0734	-	7471A	PPB	19 Sep 2017 13:12:55	U	SMPL
I5285-04 - 1	CSTP-4	67382	15.8569	-	7471A	PPB	19 Sep 2017 13:15:12	U	SMPL
I5285-05 - 1	CSTP-5	30897	7.2608	-	7471A	PPB	19 Sep 2017 13:17:28	U	SMPL
I5276-07LX5 - 1		-260	-0.0800	-	7471A	PPB	19 Sep 2017 13:19:47	U	SMPL
I5276-07A - 1		17491	4.1023	-	7471A	PPB	19 Sep 2017 13:22:05	U	SMPL
I5285-01DLX2 - 1		27357	6.4268	-	7471A	PPB	19 Sep 2017 13:24:22	U	SMPL
I5285-04DLX5 - 1		13157	3.0811	-	7471A	PPB	19 Sep 2017 13:26:39	U	SMPL
CCV58 - 1	CCV58	20224	4.7462	-	7471A	PPB	19 Sep 2017 13:28:57	U	SMPL
CCB58 - 1	CCB58	-225	-0.0717	-	7471A	PPB	19 Sep 2017 13:31:14	U	SMPL