

ISM02.4


LB90156

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44998



A= 0.0000e+000

B= 2.2261e-004

C= -1.1602e-002

Rho= 0.9999858

Accept=Accepted

glope
 $y = 1/m \cdot x + c$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.018	-0.018	-30	0.000	-30					4
0.05	0.050	0.052	0.002	284	0.0 %	284					1
0.2	0.200	0.201	0.001	956	0.0 %	956					2
2.5	2.500	2.540	0.040	11463	0.0 %	11463					-1
5.0	5.000	4.973	-0.027	22390	0.0 %	22390					0
7.5	7.500	7.498	-0.002	33733	0.0 %	33733					0
10.0	10.000	10.005	0.005	44998	0.0 %	44998					

AB
 9/19/15

LB90156
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-30	-		0 ISM02.4	PPB	9/19/2017 15:12	Std
	0.05 S0.05	284	-	0.05	ISM02.4	PPB	9/19/2017 15:14	Std
	0.2 S0.2	956	-	0.2	ISM02.4	PPB	9/19/2017 15:16	Std
	2.5 S2.5	11463	-	2.5	ISM02.4	PPB	9/19/2017 15:19	Std
	5 S5	22390	-	5	ISM02.4	PPB	9/19/2017 15:21	Std
	7.5 S7.5	33733	-	7.5	ISM02.4	PPB	9/19/2017 15:23	Std
	10 S10	44998	-	10	ISM02.4	PPB	9/19/2017 15:25	Std
ICV	ICV	17276	3.8341 -		ISM02.4	PPB	9/19/2017 15:28	SMPL
ICB	ICB	132	0.0178 -		ISM02.4	PPB	9/19/2017 15:31	SMPL
CCV059	CCV059	22416	4.9783 -		ISM02.4	PPB	9/19/2017 15:33	SMPL
CCB059	CCB059	72	0.0044 -		ISM02.4	PPB	9/19/2017 15:38	SMPL
PB102428BL	PBS004	96	0.0098 -		ISM02.4	PPB	9/19/2017 15:40	SMPL
I5240-01	MCWYL7	2430	0.5293 -		ISM02.4	PPB	9/19/2017 15:42	SMPL
I5240-02	MCWYL7D	2237	0.4864 -		ISM02.4	PPB	9/19/2017 15:45	SMPL
I5240-03	MCWYL7S	13589	3.0134 -		ISM02.4	PPB	9/19/2017 15:47	SMPL
I5240-04	MCWYL8	446	0.0877 -		ISM02.4	PPB	9/19/2017 15:49	SMPL
I5240-05	MCWYL9	2366	0.5151 -		ISM02.4	PPB	9/19/2017 15:51	SMPL
I5240-06	MCWYMO	1851	0.4004 -		ISM02.4	PPB	9/19/2017 15:54	SMPL
CCV060	CCV060	22424	4.9801 -		ISM02.4	PPB	9/19/2017 15:56	SMPL
CCB060	CCB060	83	0.0069 -		ISM02.4	PPB	9/19/2017 16:01	SMPL