


7470A

LB90179
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36985



A= 0.0000e+000

B= 2.7087e-004

C= -1.9272e-002

Rho= 0.9999854

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	28	0.000	28				
0.2	0.200	0.214	0.014	862	0.0 %	862				
2.5	2.500	2.515	0.015	9357	0.0 %	9357				
5.0	5.000	4.963	-0.037	18393	0.0 %	18393				
7.5	7.500	7.521	0.021	27836	0.0 %	27836				
10.0	10.000	9.999	-0.001	36985	0.0 %	36985				

LB90179 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Sep 2017 11:28:51

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	7470A	PPB	20 Sep 2017 11:49:06	S	Std
0.2 - 1	50.2	862	-	0.2000	7470A	PPB	20 Sep 2017 11:51:15	S	Std
2.5 - 1	52.5	9357	-	2.5000	7470A	PPB	20 Sep 2017 11:53:23	S	Std
5.0 - 1	55	18393	-	5.0000	7470A	PPB	20 Sep 2017 11:55:31	S	Std
7.5 - 1	57.5	27836	-	7.5000	7470A	PPB	20 Sep 2017 11:57:39	S	Std
10.0 - 1	510	36985	-	10.0000	7470A	PPB	20 Sep 2017 12:02:11	S	Std
ICV64 - 1	ICV64	14312	3.8574	-	7470A	PPB	20 Sep 2017 12:04:55	U	SMPL
ICB64 - 1	ICB64	-184	-0.0691	-	7470A	PPB	20 Sep 2017 12:07:01	U	SMPL
CCV61 - 1	CCV61	18238	4.9208	-	7470A	PPB	20 Sep 2017 12:09:08	U	SMPL
CCB61 - 1	CCB61	-156	-0.0615	-	7470A	PPB	20 Sep 2017 12:11:14	U	SMPL
CRA - 1	CRA	774	0.1904	-	7470A	PPB	20 Sep 2017 12:13:21	U	SMPL
HighStd - 1	HighStd	37729	10.2003	-	7470A	PPB	20 Sep 2017 12:15:28	U	SMPL
ChkStd - 1	ChkStd	24821	6.7040	-	7470A	PPB	20 Sep 2017 12:17:35	U	SMPL
PB102438BL - 1	PBW	-11	-0.0223	-	7470A	PPB	20 Sep 2017 12:19:44	U	SMPL
PB102438BS - 1	LCSW	13226	3.5632	-	7470A	PPB	20 Sep 2017 12:21:53	U	SMPL
PB102438TB - 1	PB102425TB	63	-0.0022	-	7470A	PPB	20 Sep 2017 12:24:01	U	SMPL
I5285-06 - 1	CSTP-1	119	0.0130	-	7470A	PPB	20 Sep 2017 12:26:09	U	SMPL
I5285-07 - 1	CSTP-2	171	0.0270	-	7470A	PPB	20 Sep 2017 12:28:17	U	SMPL
I5285-08 - 1	CSTP-3	399	0.0888	-	7470A	PPB	20 Sep 2017 12:30:25	U	SMPL
I5285-09 - 1	CSTP-4	114	0.0116	-	7470A	PPB	20 Sep 2017 12:32:32	U	SMPL
CCV62 - 1	CCV62	18467	4.9829	-	7470A	PPB	20 Sep 2017 12:34:38	U	SMPL
CCB62 - 1	CCB62	-142	-0.0577	-	7470A	PPB	20 Sep 2017 12:36:45	U	SMPL
I5285-10 - 1	CSTP-5	207	0.0368	-	7470A	PPB	20 Sep 2017 12:38:52	U	SMPL
I5288-02 - 1	DRUM-A-OILYSTONES	152	0.0219	-	7470A	PPB	20 Sep 2017 12:40:58	U	SMPL
I5288-04 - 1	DRUM-RAGS	121	0.0135	-	7470A	PPB	20 Sep 2017 12:43:05	U	SMPL
I5288-05 - 1	DRUM-OILY-MULCH-PAINTCHIPS	413	0.0926	-	7470A	PPB	20 Sep 2017 12:45:12	U	SMPL
I5290-04 - 1	OILY-DEBRIS	142	0.0192	-	7470A	PPB	20 Sep 2017 12:47:20	U	SMPL
I5291-02 - 1	OILY-STONE	213	0.0384	-	7470A	PPB	20 Sep 2017 12:49:27	U	SMPL
I5291-03 - 1	OILY-DEBRIS	244	0.0468	-	7470A	PPB	20 Sep 2017 12:51:35	U	SMPL
I5292-03 - 1	BU-01-091817	259	0.0509	-	7470A	PPB	20 Sep 2017 12:53:43	U	SMPL
I5292-03DUP - 1	BU-01-091817DUP	233	0.0438	-	7470A	PPB	20 Sep 2017 12:55:52	U	SMPL
I5292-03MS - 1	BU-01-091817MS	13555	3.6523	-	7470A	PPB	20 Sep 2017 12:57:59	U	SMPL
CCV63 - 1	CCV63	18338	4.9479	-	7470A	PPB	20 Sep 2017 13:00:06	U	SMPL
CCB63 - 1	CCB63	-143	-0.0580	-	7470A	PPB	20 Sep 2017 13:02:13	U	SMPL
I5292-03MSD - 1	BU-01-091817MSD	13536	3.6472	-	7470A	PPB	20 Sep 2017 13:04:20	U	SMPL
I5292-03LX5 - 1		-40	-0.0301	-	7470A	PPB	20 Sep 2017 13:06:28	U	SMPL
I5292-03A - 1		14285	3.8501	-	7470A	PPB	20 Sep 2017 13:08:34	U	SMPL
CCV64 - 1	CCV64	18028	4.8639	-	7470A	PPB	20 Sep 2017 13:10:41	U	SMPL
CCB64 - 1	CCB64	-175	-0.0667	-	7470A	PPB	20 Sep 2017 13:12:49	U	SMPL