



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90184

Temp.: 21.3 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/20/2017	13:15
2	CAL2	1	Water			7.01	09/20/2017	13:16
3	CAL3	1	Water			10.04	09/20/2017	13:17
4	ICV	1	Water			7.00	09/20/2017	13:19
5	CCV1	1	Water			2.02	09/20/2017	13:20
6	I5315-01	1	Water			4.49	09/20/2017	13:22
7	I5316-01	1	Water			4.07	09/20/2017	13:24
8	I5316-01DUP	1	Water			4.08	09/20/2017	13:25
9	CCV2	1	Water			12.01	09/20/2017	13:26

1890184

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5315 WorkList ID : 103069 Date : 9/20/2017 12:56:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5315-01-1A	pH	Cool 4 deg C	STER14	H21	JS-SYSTEM	09/17/2017	SM 4500-PH B
	Water	I5316-01-1A	pH	Cool 4 deg C	STER14	H21	MINI-CELL	09/17/2017	SM 4500-PH B

Date/Time 09.20.17 11:00 AM
Received by: JP
Relinquished by: C.B.

Date/Time 09.20.17 2:30 PM
Received by: C.B.
Relinquished by: JP