

7471A


LB90188

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42957



A= 0.0000e+000

B= 2.3444e-004

C= -1.9786e-003

Rho= 0.9999208

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	49	0.000	49				
0.2	0.200	0.233	0.033	1001	0.0 %	1001				
2.5	2.500	2.507	0.007	10700	0.0 %	10700				
5.0	5.000	4.937	-0.063	21067	0.0 %	21067				
7.5	7.500	7.445	-0.055	31765	0.0 %	31765	0			
10.0	10.000	10.069	0.069	42957	0.0 %	42957				

LB90186 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 20 Sep 2017 14:49:39					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	7471A	PPB	20 Sep 2017 14:57:22	S	Std
0.2 - 1	50.2	1001	-	0.2000	7471A	PPB	20 Sep 2017 14:59:40	S	Std
2.5 - 1	32.5	10700	-	2.5000	7471A	PPB	20 Sep 2017 15:01:57	S	Std
5.0 - 1	35	21067	-	5.0000	7471A	PPB	20 Sep 2017 15:04:15	S	Std
7.5 - 1	37.5	31765	-	7.5000	7471A	PPB	20 Sep 2017 15:06:32	S	Std
10.0 - 1	510	42957	-	10.0000	7471A	PPB	20 Sep 2017 15:13:58	S	Std
ICV65 - 1	ICV65	16249	3.8075	-	7471A	PPB	20 Sep 2017 15:16:36	U	SMPL
ICB65 - 1	ICB65	-168	-0.0414	-	7471A	PPB	20 Sep 2017 15:18:52	U	SMPL
CCV65 - 1	CCV65	21447	5.0261	-	7471A	PPB	20 Sep 2017 15:21:08	U	SMPL
CCB65 - 1	CCB65	-175	-0.0430	-	7471A	PPB	20 Sep 2017 15:23:24	U	SMPL
CRA - 1	CRA	943	0.2191	-	7471A	PPB	20 Sep 2017 15:25:40	U	SMPL
HighStd - 1	HighStd	43520	10.2010	-	7471A	PPB	20 Sep 2017 15:27:56	U	SMPL
ChkStd - 1	ChkStd	28434	6.6642	-	7471A	PPB	20 Sep 2017 15:30:12	U	SMPL
PB102450BL - 1	PBS	-207	-0.0505	-	7471A	PPB	20 Sep 2017 15:32:30	U	SMPL
PB102450BS - 1	LCSS	15596	3.6544	-	7471A	PPB	20 Sep 2017 15:34:47	U	SMPL
I5288-01 - 1	DRUM-A-OILYSTONES	-48	-0.0132	-	7471A	PPB	20 Sep 2017 15:37:04	U	SMPL
I5290-02 - 1	OILY-DEBRIS	49	0.0095	-	7471A	PPB	20 Sep 2017 15:39:22	U	SMPL
I5291-01 - 1	OILY-STONE	-25	-0.0078	-	7471A	PPB	20 Sep 2017 15:41:39	U	SMPL
I5292-01 - 1	BU-01-091817	640	0.1481	-	7471A	PPB	20 Sep 2017 15:43:57	U	SMPL
I5293-01 - 1	TP-2-COMP	10140	2.3753	-	7471A	PPB	20 Sep 2017 15:46:13	U	SMPL
CCV66 - 1	CCV66	21469	5.0313	-	7471A	PPB	20 Sep 2017 15:48:29	U	SMPL
CCB66 - 1	CCB66	-195	-0.0477	-	7471A	PPB	20 Sep 2017 15:50:45	U	SMPL
I5293-04 - 1	TP-1-COMP	4311	1.0087	-	7471A	PPB	20 Sep 2017 15:53:01	U	SMPL
I5293-07 - 1	SP-A-COMP	5331	1.2478	-	7471A	PPB	20 Sep 2017 15:55:17	U	SMPL
I5293-10 - 1	SP-B-R-SIDE-COMP	11933	2.7956	-	7471A	PPB	20 Sep 2017 15:57:33	U	SMPL
I5293-13 - 1	SP-B-L-SIDE-COMP	4592	1.0746	-	7471A	PPB	20 Sep 2017 15:59:49	U	SMPL
I5293-13DUP - 1	SP-B-L-SIDE-COMPDUP	5118	1.1979	-	7471A	PPB	20 Sep 2017 16:02:06	U	SMPL
I5293-13MS - 1	SP-B-L-SIDE-COMPMS	19934	4.6714	-	7471A	PPB	20 Sep 2017 16:04:23	U	SMPL
I5293-13MSD - 1	SP-B-L-SIDE-COMPMSD	21691	5.0833	-	7471A	PPB	20 Sep 2017 16:06:42	U	SMPL
I5304-01 - 1	SASP2P-ENV-117-2(0-5)	878	0.2039	-	7471A	PPB	20 Sep 2017 16:09:01	U	SMPL
I5304-02 - 1	SASP2P-ENV-117-2(5-10)	112	0.0243	-	7471A	PPB	20 Sep 2017 16:11:18	U	SMPL
I5304-03 - 1	SASP2P-ENV-116-4(30-40)	17	0.0020	-	7471A	PPB	20 Sep 2017 16:13:34	U	SMPL
CCV67 - 1	CCV67	21526	5.0447	-	7471A	PPB	20 Sep 2017 16:15:51	U	SMPL
CCB67 - 1	CCB67	-147	-0.0364	-	7471A	PPB	20 Sep 2017 16:18:07	U	SMPL
I5304-04 - 1	SASP2P-ENV-118-4(0-5)	6583	1.5414	-	7471A	PPB	20 Sep 2017 16:20:24	U	SMPL
I5304-05 - 1	SASP2P-ENV-118-4(5-10)	7415	1.7364	-	7471A	PPB	20 Sep 2017 16:22:40	U	SMPL
I5306-01 - 1	EP-21	252	0.0571	-	7471A	PPB	20 Sep 2017 16:24:56	U	SMPL
I5311-01 - 1	OK-03-091917	2475	0.5783	-	7471A	PPB	20 Sep 2017 16:27:13	U	SMPL
I5312-01 - 1	OR-01-091917	2199	0.5136	-	7471A	PPB	20 Sep 2017 16:29:29	U	SMPL
I5313-01 - 1	72-12000-R4-COMP	1172	0.2728	-	7471A	PPB	20 Sep 2017 16:31:46	U	SMPL
I5293-13LX5 - 1		900	0.2090	-	7471A	PPB	20 Sep 2017 16:34:04	U	SMPL
I5293-13A - 1		21879	5.1274	-	7471A	PPB	20 Sep 2017 16:36:21	U	SMPL
CCV68 - 1	CCV68	21486	5.0353	-	7471A	PPB	20 Sep 2017 16:38:39	U	SMPL
CCB68 - 1	CCB68	-200	-0.0489	-	7471A	PPB	20 Sep 2017 16:40:55	U	SMPL