



Analytical Summary Report

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB90188

Reagent/Standard	Lot/Log #
conductivity standard	WP54988

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB90188BL		1	21.3	0.92	0.99	09/20/2017	14:15
2	LB90188BS	1413	1	20.8	1261.00	1370.98	09/20/2017	14:24
3	I5274-08		1	21.1	257.00	277.68	09/20/2017	14:37

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$