



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90201

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/20/2017	17:10
2	CAL2	1	Water			7.00	09/20/2017	17:11
3	CAL3	1	Water			10.02	09/20/2017	17:12
4	ICV	1	Water			7.00	09/20/2017	17:13
5	CCV1	1	Water			2.01	09/20/2017	17:15
6	I5309-04	1	Solid	20.02	20	5.87	09/20/2017	17:16
7	I5309-08	1	Solid	20.03	20	6.02	09/20/2017	17:18
8	I5329-03	1	Solid	20.01	20	8.81	09/20/2017	17:20
9	I5330-03	1	Solid	20.03	20	8.46	09/20/2017	17:22
10	I5331-03	1	Solid	20.05	20	9.05	09/20/2017	17:24
11	I5331-06	1	Solid	20.04	20	9.14	09/20/2017	17:25
12	I5331-09	1	Solid	20.01	20	9.07	09/20/2017	17:26
13	I5335-02	1	Solid	20.03	20	8.00	09/20/2017	17:28
14	I5336-03	1	Solid	20.02	20	8.20	09/20/2017	17:30
15	I5336-03DUP	1	Solid	20.03	20	8.21	09/20/2017	17:31
16	CCV2	1	Water			12.01	09/20/2017	17:32

LB96201

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5332

WorkList ID : 103080

Date : 9/20/2017 5:53:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/22/2017	Solid	I5309-04	Corrosivity	Cool 4 deg C	PSEG01	H32	TP-1-4-COMP	09/19/2017	9045C
09/22/2017	Solid	I5309-08	Corrosivity	Cool 4 deg C	PSEG01	H32	TP-5-COMP	09/19/2017	9045C
	Solid	I5329-03	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-01092017-COMP	09/20/2017	9045C
	Solid	I5330-03	Corrosivity	Cool 4 deg C	PSEG05	H52	HD-01092017-COMP	09/20/2017	9045C
	Solid	I5331-03	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-05-92017-A	09/20/2017	9040C
	Solid	I5331-06	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-05-92017-B	09/20/2017	9040C
	Solid	I5331-09	Corrosivity	Cool 4 deg C	PSEG05	H52	HR-05-92017-C	09/20/2017	9040C
	Solid	I5335-02	Corrosivity	Cool 4 deg C	PSEG01	H52	SP-COMP	09/20/2017	9045C
	Solid	I5336-03	Corrosivity	Cool 4 deg C	PSEG05	H53	TR-05-092017	09/20/2017	9045C

Date/Time 09-20-17 6:00p
 Received by: JP
 Relinquished by: JP

Date/Time 09-20-17 2:55p
 Received by: W
 Relinquished by: JP