



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90202

Temp.: 22.2 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/20/2017	17:35
2	CAL2	1	Water			7.01	09/20/2017	17:36
3	CAL3	1	Water			10.03	09/20/2017	17:37
4	ICV	1	Water			6.99	09/20/2017	17:38
5	CCV1	1	Water			2.02	09/20/2017	17:40
6	I5333-02	1	Water			7.68	09/20/2017	17:41
7	I5339-02	1	Water			7.98	09/20/2017	17:42
8	I5339-02DUP	1	Water			7.99	09/20/2017	17:43
9	CCV2	1	Water			12.01	09/20/2017	17:44

1390202

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I5333 WorkList ID : 103081 Date : 9/20/2017 5:53:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5333-02 -1	Corrosivity	Cool 4 deg C	PSEG01	H52	SV-30745L	09/20/2017	9040C
	Water	I5339-02 -1	Corrosivity	Cool 4 deg C	PSEG01	H51	FRAC-TANK-573094	09/20/2017	9040C

Date/Time 09.20.17 6:00 PM
Received by: JP
Relinquished by: JR

Date/Time 09.20.17 7:20 PM
Received by:
Relinquished by: JP