



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90207

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP58109
Turbidity std, 40NTU	WP58103
turbidity std 5 NTU	WP58106
turbidity std 20 NTU	WP58104
Turbidity std 1 NTU	WP58107
turbidity std 0 NTU	WP58110
turbidity std 10 NTU	WP58105

Intercept: 0.0131

Slope: 0.9955

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.094	09/21/2017	10:16
2	CAL2	1	1	1.031	09/21/2017	10:17
3	CAL3	5	1	4.950	09/21/2017	10:19
4	CAL4	10	1	9.960	09/21/2017	10:22
5	CAL5	20	1	19.800	09/21/2017	10:24
6	CAL6	40	1	39.900	09/21/2017	10:27



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90207

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.700	9.731	09/21/2017	10:30
2	ICB		1	0.092	0.079	09/21/2017	10:33
3	CCV1	10	1	9.400	9.429	09/21/2017	12:15
4	CCB1		1	0.099	0.086	09/21/2017	12:18
5	LB90207BL		1	0.095	0.082	09/21/2017	12:21
6	I5274-37		1	10.100	10.132	09/21/2017	12:23
7	CCV2	10	1	9.400	9.429	09/21/2017	12:25
8	CCB2		1	0.089	0.076	09/21/2017	12:28