



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90208

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP58108
TURBIDITY CCV 10 NTU	WP58109
Turbidity std, 40NTU	WP58103
turbidity std 5 NTU	WP58106
turbidity std 20 NTU	WP58104
Turbidity std 1 NTU	WP58107
turbidity std 0 NTU	WP58110
turbidity std 10 NTU	WP58105

Intercept: 0.0131

Slope: 0.9955

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.094	09/21/2017	10:16
2	CAL2	1	1	1.031	09/21/2017	10:17
3	CAL3	5	1	4.950	09/21/2017	10:19
4	CAL4	10	1	9.960	09/21/2017	10:22
5	CAL5	20	1	19.800	09/21/2017	10:24
6	CAL6	40	1	39.900	09/21/2017	10:27



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90208

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.700	9.731	09/21/2017	10:30
2	ICB		1	0.092	0.079	09/21/2017	10:33
3	CCV1	10	1	9.410	9.439	09/21/2017	11:53
4	CCB1		1	0.099	0.086	09/21/2017	11:56
5	LB90208BL		1	0.097	0.084	09/21/2017	12:03
6	I5274-36		1	10.200	10.233	09/21/2017	12:06
7	CCV2	10	1	9.560	9.590	09/21/2017	12:09
8	CCB2		1	0.094	0.081	09/21/2017	12:12