


7470A

LB90209
INSTRUMENT ID: CY1

Linear

 μ Abs.:

36823



A= 0.0000e+000

B= 2.6916e-004

C= -1.3639e-003

Rho= 0.9998425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-29	0.000	-29				
0.2	0.200	0.210	0.010	786	0.0 %	786				
2.5	2.500	2.440	-0.060	9071	0.0 %	9071				
5.0	5.000	5.038	0.038	18721	0.0 %	18721				
7.5	7.500	7.611	0.111	28284	0.0 %	28284				
10.0	10.000	9.910	-0.090	36823	0.0 %	36823				

LB90209 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Sep 2017 09:39:20				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-29	-	0.0000	7470A	PPB	21 Sep 2017 10:34:41	S Std
0.2 - 1	50.2	786	-	0.2000	7470A	PPB	21 Sep 2017 10:36:50	S Std
2.5 - 1	52.5	9071	-	2.5000	7470A	PPB	21 Sep 2017 10:38:58	S Std
5.0 - 1	55	18721	-	5.0000	7470A	PPB	21 Sep 2017 10:41:06	S Std
7.5 - 1	57.5	28284	-	7.5000	7470A	PPB	21 Sep 2017 10:43:14	S Std
10.0 - 1	570	36823	-	10.0000	7470A	PPB	21 Sep 2017 10:47:49	S Std
ICV66 - 1	ICV66	14670	3.9472	-	7470A	PPB	21 Sep 2017 10:50:27	U SMPL
ICB66 - 1	ICB66	-130	-0.0364	-	7470A	PPB	21 Sep 2017 10:52:34	U SMPL
CCV69 - 1	CCV69	17416	4.6863	-	7470A	PPB	21 Sep 2017 10:54:41	U SMPL
CCB69 - 1	CCB69	-172	-0.0477	-	7470A	PPB	21 Sep 2017 10:56:48	U SMPL
CRA - 1	CRA	656	0.1752	-	7470A	PPB	21 Sep 2017 10:58:55	U SMPL
HighStd - 1	HighStd	36753	9.8910	-	7470A	PPB	21 Sep 2017 11:01:02	U SMPL
ChkStd - 1	ChkStd	26347	7.0901	-	7470A	PPB	21 Sep 2017 11:03:09	U SMPL
PB102479BL - 1	PBW	-271	-0.0743	-	7470A	PPB	21 Sep 2017 11:05:17	U SMPL
PB102479BS - 1	LCSW	14908	4.0112	-	7470A	PPB	21 Sep 2017 11:09:50	U SMPL
PB102479TB - 1	PB102439TB	-211	-0.0582	-	7470A	PPB	21 Sep 2017 11:11:58	U SMPL
I5293-03 - 1	TP-2-COMP	-158	-0.0439	-	7470A	PPB	21 Sep 2017 11:14:06	U SMPL
I5293-06 - 1	TP-1-COMP	-126	-0.0353	-	7470A	PPB	21 Sep 2017 11:16:14	U SMPL
I5293-09 - 1	SP-A-COMP	-77	-0.0221	-	7470A	PPB	21 Sep 2017 11:18:23	U SMPL
I5293-12 - 1	SP-B-R-SIDE-COMP	-116	-0.0326	-	7470A	PPB	21 Sep 2017 11:20:30	U SMPL
CCV70 - 1	CCV70	18406	4.9527	-	7470A	PPB	21 Sep 2017 11:22:37	U SMPL
CCB70 - 1	CCB70	-115	-0.0323	-	7470A	PPB	21 Sep 2017 11:24:43	U SMPL
I5293-15 - 1	SP-B-L-SIDE-COMP	-74	-0.0213	-	7470A	PPB	21 Sep 2017 11:26:50	U SMPL
I5308-04 - 1	220-1-COMP	-113	-0.0318	-	7470A	PPB	21 Sep 2017 11:28:57	U SMPL
I5308-08 - 1	220-1(O-U)-COMP	-99	-0.0280	-	7470A	PPB	21 Sep 2017 11:31:04	U SMPL
I5310-03 - 1	NB-01-091917-A	-119	-0.0334	-	7470A	PPB	21 Sep 2017 11:33:11	U SMPL
I5310-06 - 1	NB-01-091917-B	-107	-0.0302	-	7470A	PPB	21 Sep 2017 11:35:19	U SMPL
I5311-03 - 1	OK-03-091917	-80	-0.0229	-	7470A	PPB	21 Sep 2017 11:37:27	U SMPL
I5312-03 - 1	OR-01-091917	-89	-0.0253	-	7470A	PPB	21 Sep 2017 11:39:35	U SMPL
I5313-06 - 1	72-12000-R4-COMP	-75	-0.0216	-	7470A	PPB	21 Sep 2017 11:41:43	U SMPL
I5313-06DUP - 1	72-12000-R4-COMP	-76	-0.0218	-	7470A	PPB	21 Sep 2017 11:43:52	U SMPL
I5313-06MS - 1	72-12000-R4-COMPMS	14575	3.9216	-	7470A	PPB	21 Sep 2017 11:45:59	U SMPL
CCV71 - 1	CCV71	18388	4.9479	-	7470A	PPB	21 Sep 2017 11:48:05	U SMPL
CCB71 - 1	CCB71	-165	-0.0458	-	7470A	PPB	21 Sep 2017 11:50:12	U SMPL
I5313-06MSD - 1	72-12000-R4-COMPMSD	13190	3.5488	-	7470A	PPB	21 Sep 2017 11:52:20	U SMPL
PB102495BL - 1	PBW	-200	-0.0552	-	7470A	PPB	21 Sep 2017 11:54:27	U SMPL
PB102495BS - 1	LCSW	14233	3.8296	-	7470A	PPB	21 Sep 2017 11:56:34	U SMPL
I5301-01 - 1	20170605-COMP-B	-279	-0.0765	-	7470A	PPB	21 Sep 2017 11:58:41	U SMPL
I5301-01DUP - 1	20170605-COMP-BDUP	-204	-0.0563	-	7470A	PPB	21 Sep 2017 12:00:49	U SMPL
I5301-01MS - 1	20170605-COMP-BMS	14252	3.8347	-	7470A	PPB	21 Sep 2017 12:08:26	U SMPL
I5301-01MSD - 1	20170605-COMP-BMSD	13959	3.7558	-	7470A	PPB	21 Sep 2017 12:10:34	U SMPL
I5301-02 - 1	20170607-COMP-D	-306	-0.0837	-	7470A	PPB	21 Sep 2017 12:12:42	U SMPL
I5301-03 - 1	20170604-COMP-A	-167	-0.0463	-	7470A	PPB	21 Sep 2017 12:14:50	U SMPL
I5301-04 - 1	20170606-COMP-C	-160	-0.0444	-	7470A	PPB	21 Sep 2017 12:16:58	U SMPL
CCV72 - 1	CCV72	18349	4.9374	-	7470A	PPB	21 Sep 2017 12:19:05	U SMPL
CCB72 - 1	CCB72	-151	-0.0420	-	7470A	PPB	21 Sep 2017 12:21:12	U SMPL
I5301-05 - 1	20170608-COMP-A	-154	-0.0428	-	7470A	PPB	21 Sep 2017 12:23:19	U SMPL
I5301-06 - 1	20170609-COMP-A	-167	-0.0463	-	7470A	PPB	21 Sep 2017 12:25:27	U SMPL
I5301-08 - 1	20170605-COMP-B	-174	-0.0482	-	7470A	PPB	21 Sep 2017 12:27:34	U SMPL
I5301-09 - 1	20170607-COMP-D	-150	-0.0417	-	7470A	PPB	21 Sep 2017 12:29:41	U SMPL
I5301-10 - 1	20170604-COMP-A	-170	-0.0471	-	7470A	PPB	21 Sep 2017 12:31:49	U SMPL
I5301-11 - 1	20170606-COMP-C	-152	-0.0423	-	7470A	PPB	21 Sep 2017 12:33:56	U SMPL
I5301-12 - 1	20170608-COMP-A	-156	-0.0434	-	7470A	PPB	21 Sep 2017 12:36:05	U SMPL
I5301-13 - 1	20170609-COMP-A	-176	-0.0487	-	7470A	PPB	21 Sep 2017 12:38:13	U SMPL
CCV73 - 1	CCV73	17975	4.8367	-	7470A	PPB	21 Sep 2017 12:40:21	U SMPL
CCB73 - 1	CCB73	-198	-0.0547	-	7470A	PPB	21 Sep 2017 12:42:28	U SMPL
I5313-06LX5 - 1		-105	-0.0296	-	7470A	PPB	21 Sep 2017 12:44:36	U SMPL
I5313-06A - 1		14549	3.9146	-	7470A	PPB	21 Sep 2017 12:46:43	U SMPL
I5301-01LX5 - 1		-205	-0.0565	-	7470A	PPB	21 Sep 2017 12:48:51	U SMPL
I5301-01A - 1		15786	4.2476	-	7470A	PPB	21 Sep 2017 12:50:58	U SMPL
CCV74 - 1	CCV74	17970	4.8354	-	7470A	PPB	21 Sep 2017 12:53:05	U SMPL
CCB74 - 1	CCB74	-211	-0.0582	-	7470A	PPB	21 Sep 2017 12:55:12	U SMPL