

LB90210

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

9/20/2017 18:32

Test: Phenolics-

mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.008	0.0	0.113	
ICB1	0.005	0.0	0.003	
CCV1	1.022	0.0	0.114	
CCB1	-0.014	0.0	0.000	
PB102422BL	0.006	0.0	0.003	
PB10242BS	1.013	0.0	0.113	
I5274-26	1.634	0.0	0.182	
CCV2	1.002	0.0	0.112	
CCB2	0.005	0.0	0.003	

N	9
Mean	0.631
SD	0.6299
CV%	99.79

LB90210

Aquakem v. 7.2AQ1

Results from time period:

Wed Sep 20 18:00:53 2017

Wed Sep 20 18:01:01 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.056	mg/l	9/20/2017 16:41:58	
0.1PPM	A	Phenolics-	P	0.1013	mg/l	9/20/2017 16:41:59	
0.25PPM	A	Phenolics-	P	0.2568	mg/l	9/20/2017 16:42:00	
0.50PPM	A	Phenolics-	P	0.4993	mg/l	9/20/2017 16:42:01	
1.0PPM	A	Phenolics-	P	0.9752	mg/l	9/20/2017 16:42:02	
2.0PPM	A	Phenolics-	P	2.0115	mg/l	9/20/2017 16:42:03	
ICV1	S	Phenolics-	P	1.008	mg/l	9/20/2017 18:00:53	
ICB1	S	Phenolics-	P	0.0053	mg/l	9/20/2017 18:00:54	
CCV1	S	Phenolics-	P	1.022	mg/l	9/20/2017 18:00:55	
CCB1	S	Phenolics-	P	-0.0143	mg/l	9/20/2017 18:00:56	
PB102422BL	S	Phenolics-	P	0.0059	mg/l	9/20/2017 18:00:57	
PB10242BS	S	Phenolics-	P	1.0134	mg/l	9/20/2017 18:00:58	
I5274-26	S	Phenolics-	P	1.6335	mg/l	9/20/2017 18:00:59	
CCV2	S	Phenolics-	P	1.0024	mg/l	9/20/2017 18:01:00	
CCB2	S	Phenolics-	P	0.0052	mg/l	9/20/2017 18:01:01	

LB90210

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

9/20/2017 16:44


Test Phenolics-

Accepted 9/20/2017 16:44

Factor 9.1
Bias 0.002

Coeff. of det. 0.999703

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0560	0.0500	
2	Phenol-2	0.013	0.1013	0.1000	
3	Phenol-2	0.030	0.2568	0.2500	
4	Phenol-2	0.057	0.4993	0.5000	
5	Phenol-2	0.109	0.9752	1.0000	
6	Phenol-2	0.223	2.0115	2.0000	