



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB90214

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP58116
COD calibration std. 100 ppm	WP58113
COD calibration std. 150 ppm	WP58114
COD calibration std. 50 ppm	WP58115
COD calibration std. 0 ppm	WP58111
COD CCV std, 50ppm	WP58117
COD Digestion Vials Low Level 0-150Mg/L	W2217

Temp In (C) : <u>150</u>	Date In: <u>09/21/2017</u>	Time In: <u>09:40</u>
Temp Out (C) : <u>150</u>	Date Out: <u>09/21/2017</u>	Time Out: <u>11:40</u>

Intercept: 0.4899

Slope: 1.0276

Regression: 0.9997

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/21/2017	12:44
2	CAL2	10	1	Water	12.000	09/21/2017	12:45
3	CAL3	50	1	Water	50.000	09/21/2017	12:45
4	CAL4	100	1	Water	105.000	09/21/2017	12:46
5	CAL5	150	1	Water	154.000	09/21/2017	12:46



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB90214

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	48.180	09/21/2017	12:47
2	ICB				1	Water	0.000	-0.477	09/21/2017	12:47
3	CCV1	50			1	Water	51.000	49.153	09/21/2017	12:48
4	CCB1				1	Water	0.000	-0.477	09/21/2017	12:48
5	LB90214BL				1	Water	0.000	-0.477	09/21/2017	12:49
6	LB90214BS	50			1	Water	48.000	46.234	09/21/2017	12:49
7	I5274-03				2	Water	101.000	97.811	09/21/2017	12:50
8	I5322-02				1	Water	5.000	4.389	09/21/2017	12:50
9	I5322-02DUP				1	Water	5.000	4.389	09/21/2017	12:51
10	I5322-02MS	50			1	Water	55.000	53.046	09/21/2017	12:51
11	I5322-02MSD	50			1	Water	56.000	54.019	09/21/2017	12:52
12	I5322-03				1	Water	0.000	-0.477	09/21/2017	12:52
13	I5333-01				1	Water	3.000	2.443	09/21/2017	12:53
14	I5339-01				1	Water	11.000	10.228	09/21/2017	12:53
15	CCV2	50			1	Water	51.000	49.153	09/21/2017	12:54
16	CCB2				1	Water	0.000	-0.477	09/21/2017	12:54