


245.1

LB90221
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32290



A= 0.0000e+000

B= 3.0906e-004

C= -7.5628e-004

Rho= 0.9999711

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-62	0.000	-62				
0.2	0.200	0.173	-0.027	561	0.0 %	561				
2.5	2.500	2.555	0.055	8269	0.0 %	8269				
5.0	5.000	5.011	0.011	16215	0.0 %	16215				
7.5	7.500	7.503	0.003	24280	0.0 %	24280				
10.0	10.000	9.979	-0.021	32290	0.0 %	32290				

LB90221 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Sep 2017 14:14:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-62	-	0.0000	245.1	PPB	21 Sep 2017 14:19:05	S	Std
0.2 - 1	30.2	561	-	0.2000	245.1	PPB	21 Sep 2017 14:21:13	S	Std
2.5 - 1	52.5	8269	-	2.5000	245.1	PPB	21 Sep 2017 14:23:22	S	Std
5.0 - 1	55	16215	-	5.0000	245.1	PPB	21 Sep 2017 14:25:30	S	Std
7.5 - 1	57.5	24280	-	7.5000	245.1	PPB	21 Sep 2017 14:27:38	S	Std
10.0 - 1	510	32290	-	10.0000	245.1	PPB	21 Sep 2017 14:29:46	S	Std
ICV67 - 1	ICV67	12572	3.8847	-245.1		PPB	21 Sep 2017 14:37:29	U	SMPL
ICB67 - 1	ICB67	-165	-0.0518	-245.1		PPB	21 Sep 2017 14:39:35	U	SMPL
CCV75 - 1	CCV75	16283	5.0316	-245.1		PPB	21 Sep 2017 14:41:42	U	SMPL
CCB75 - 1	CCB75	-206	-0.0644	-245.1		PPB	21 Sep 2017 14:43:49	U	SMPL
CRA - 1	CRA	554	0.1705	-245.1		PPB	21 Sep 2017 14:45:56	U	SMPL
HighStd - 1	HighStd	32154	9.9367	-245.1		PPB	21 Sep 2017 14:48:03	U	SMPL
ChkStd - 1	ChkStd	22472	6.9444	-245.1		PPB	21 Sep 2017 14:50:09	U	SMPL
PB102478BL - 1	PBW	-276	-0.0861	-245.1		PPB	21 Sep 2017 14:52:18	U	SMPL
PB102478BS - 1	LCSW	12364	3.8204	-245.1		PPB	21 Sep 2017 14:54:27	U	SMPL
I5297-01 - 1	091517-SIDI-E-D-B	-281	-0.0876	-245.1		PPB	21 Sep 2017 14:56:34	U	SMPL
I5297-02 - 1	091517-SIDI-E-D-S	-220	-0.0687	-245.1		PPB	21 Sep 2017 14:58:42	U	SMPL
I5297-03 - 1	091517-SIDI-E-U-B	-174	-0.0545	-245.1		PPB	21 Sep 2017 15:00:51	U	SMPL
I5297-04 - 1	091517-SIDI-E-U-S	-185	-0.0579	-245.1		PPB	21 Sep 2017 15:02:59	U	SMPL
I5297-05 - 1	091517-TIDO-E-D-M	-193	-0.0604	-245.1		PPB	21 Sep 2017 15:05:06	U	SMPL
CCV76 - 1	CCV76	16246	5.0202	-245.1		PPB	21 Sep 2017 15:07:12	U	SMPL
CCB76 - 1	CCB76	-185	-0.0579	-245.1		PPB	21 Sep 2017 15:09:19	U	SMPL
I5297-06 - 1	091517-TIDO-E-U-M	-201	-0.0629	-245.1		PPB	21 Sep 2017 15:11:26	U	SMPL
I5298-01 - 1	091817-SIDI-F-D-B	-182	-0.0570	-245.1		PPB	21 Sep 2017 15:13:33	U	SMPL
I5298-02 - 1	091817-SIDI-F-D-M	-157	-0.0493	-245.1		PPB	21 Sep 2017 15:15:39	U	SMPL
I5298-03 - 1	091817-SIDI-F-D-S	-128	-0.0403	-245.1		PPB	21 Sep 2017 15:17:46	U	SMPL
I5298-04 - 1	091817-SIDI-F-U-B	-224	-0.0700	-245.1		PPB	21 Sep 2017 15:19:54	U	SMPL
I5298-05 - 1	091817-SIDI-F-U-M	-147	-0.0462	-245.1		PPB	21 Sep 2017 15:22:04	U	SMPL
I5298-06 - 1	091817-SIDI-F-U-S	-176	-0.0552	-245.1		PPB	21 Sep 2017 15:24:12	U	SMPL
I5298-06DUP - 1	091817-SIDI-F-U-SDUP	-180	-0.0564	-245.1		PPB	21 Sep 2017 15:26:20	U	SMPL
I5298-06MS - 1	091817-SIDI-F-U-SMS	12452	3.8476	-245.1		PPB	21 Sep 2017 15:31:38	U	SMPL
I5298-06MSD - 1	091817-SIDI-F-U-SMSD	12271	3.7917	-245.1		PPB	21 Sep 2017 15:33:45	U	SMPL
CCV77 - 1	CCV77	16369	5.0582	-245.1		PPB	21 Sep 2017 15:35:52	U	SMPL
CCB77 - 1	CCB77	-201	-0.0629	-245.1		PPB	21 Sep 2017 15:37:59	U	SMPL
I5298-06LX5 - 1		-119	-0.0375	-245.1		PPB	21 Sep 2017 15:40:07	U	SMPL
I5298-06A - 1		12685	3.9196	-245.1		PPB	21 Sep 2017 15:44:47	U	SMPL
CCV78 - 1	CCV78	16154	4.9918	-245.1		PPB	21 Sep 2017 15:46:53	U	SMPL
CCB78 - 1	CCB78	-194	-0.0607	-245.1		PPB	21 Sep 2017 15:49:00	U	SMPL