


7471A

LB90226
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42147



A= 0.0000e+000

B= 2.3819e-004

C= 1.3894e-003

Rho= 0.9999434

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	10	0.000	10				
0.2	0.200	0.191	-0.009	798	0.0 %	798				
2.5	2.500	2.513	0.013	10546	0.0 %	10546				
5.0	5.000	5.031	0.031	21116	0.0 %	21116				
7.5	7.500	7.420	-0.080	31146	0.0 %	31146	0			
10.0	10.000	10.040	0.040	42147	0.0 %	42147				

LB90226 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Sep 2017 15:58:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471A	PPB	21 Sep 2017 16:02:33	S	Std
0.2 - 1	50.2	798	-	0.2000	7471A	PPB	21 Sep 2017 16:04:50	S	Std
2.5 - 1	52.5	10546	-	2.5000	7471A	PPB	21 Sep 2017 16:07:08	S	Std
5.0 - 1	55	21116	-	5.0000	7471A	PPB	21 Sep 2017 16:09:25	S	Std
7.5 - 1	57.5	31146	-	7.5000	7471A	PPB	21 Sep 2017 16:11:42	S	Std
10.0 - 1	510	42147	-	10.0000	7471A	PPB	21 Sep 2017 16:13:59	S	Std
ICV68 - 1	ICV68	16620	3.9601	-	7471A	PPB	21 Sep 2017 16:19:37	U	SMPL
ICB68 - 1	ICB68	-187	-0.0432	-	7471A	PPB	21 Sep 2017 16:21:53	U	SMPL
CCV79 - 1	CCV79	21256	5.0644	-	7471A	PPB	21 Sep 2017 16:24:09	U	SMPL
CCB79 - 1	CCB79	-185	-0.0427	-	7471A	PPB	21 Sep 2017 16:26:24	U	SMPL
CRA - 1	CRA	761	0.1827	-	7471A	PPB	21 Sep 2017 16:28:41	U	SMPL
HighStd - 1	HighStd	42195	10.0518	-	7471A	PPB	21 Sep 2017 16:30:57	U	SMPL
ChkStd - 1	ChkStd	28127	6.7010	-	7471A	PPB	21 Sep 2017 16:33:13	U	SMPL
PB102485BL - 1	PBS	-330	-0.0772	-	7471A	PPB	21 Sep 2017 16:35:31	U	SMPL
PB102485BS - 1	LCSS	15122	3.6033	-	7471A	PPB	21 Sep 2017 16:40:20	U	SMPL
I5308-01 - 1	220-1-COMP	30308	7.2204	-	7471A	PPB	21 Sep 2017 16:42:37	U	SMPL
I5308-05 - 1	220-1-(O-U)-COMP	6395	1.5246	-	7471A	PPB	21 Sep 2017 16:44:54	U	SMPL
I5309-01 - 1	TP-1-4-COMP	1719	0.4108	-	7471A	PPB	21 Sep 2017 16:52:07	U	SMPL
I5309-01DUP - 1	TP-1-4-COMPDUP	1715	0.4099	-	7471A	PPB	21 Sep 2017 16:54:24	U	SMPL
I5309-01MS - 1	TP-1-4-COMPMS	15384	3.6657	-	7471A	PPB	21 Sep 2017 16:56:40	U	SMPL
CCV80 - 1	CCV80	21317	5.0789	-	7471A	PPB	21 Sep 2017 16:58:57	U	SMPL
CCB80 - 1	CCB80	-199	-0.0460	-	7471A	PPB	21 Sep 2017 17:01:13	U	SMPL
I5309-01MSD - 1	TP-1-4-COMPMSD	14842	3.5366	-	7471A	PPB	21 Sep 2017 17:03:28	U	SMPL
I5309-05 - 1	TP-5-COMP	1057	0.2532	-	7471A	PPB	21 Sep 2017 17:05:44	U	SMPL
I5310-01 - 1	NB-01-091917-A	967	0.2317	-	7471A	PPB	21 Sep 2017 17:08:00	U	SMPL
I5310-04 - 1	NB-01-091917-B	816	0.1958	-	7471A	PPB	21 Sep 2017 17:10:17	U	SMPL
I5329-01 - 1	CL-01092017-COMP	2090	0.4992	-	7471A	PPB	21 Sep 2017 17:12:34	U	SMPL
I5330-01 - 1	HD-01092017-COMP	916	0.2196	-	7471A	PPB	21 Sep 2017 17:14:51	U	SMPL
I5331-01 - 1	HR-05-92017-A	4990	1.1900	-	7471A	PPB	21 Sep 2017 17:17:08	U	SMPL
I5331-04 - 1	HR-05-92017-B	5679	1.3541	-	7471A	PPB	21 Sep 2017 17:19:26	U	SMPL
I5331-07 - 1	HR-05-92017-C	5647	1.3464	-	7471A	PPB	21 Sep 2017 17:21:43	U	SMPL
I5335-01 - 1	SP-COMP	5976	1.4248	-	7471A	PPB	21 Sep 2017 17:23:59	U	SMPL
CCV81 - 1	CCV81	21035	5.0117	-	7471A	PPB	21 Sep 2017 17:26:16	U	SMPL
CCB81 - 1	CCB81	-154	-0.0353	-	7471A	PPB	21 Sep 2017 17:28:31	U	SMPL
I5336-01 - 1	TR-05-092017	1600	0.3825	-	7471A	PPB	21 Sep 2017 17:30:48	U	SMPL
I5309-01LX5 - 1		304	0.0738	-	7471A	PPB	21 Sep 2017 17:33:05	U	SMPL
I5309-01A - 1		18478	4.4027	-	7471A	PPB	21 Sep 2017 17:35:21	U	SMPL
CCV82 - 1	CCV82	21153	5.0398	-	7471A	PPB	21 Sep 2017 17:43:09	U	SMPL
CCB82 - 1	CCB82	-156	-0.0358	-	7471A	PPB	21 Sep 2017 17:45:25	U	SMPL