



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90237

Temp.: 21.2 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/21/2017	16:45
2	CAL2	1	Water			7.00	09/21/2017	16:46
3	CAL3	1	Water			10.04	09/21/2017	16:47
4	ICV	1	Water			6.99	09/21/2017	16:48
5	CCV1	1	Water			2.02	09/21/2017	16:50
6	I5361-03	1	Water			6.50	09/21/2017	16:51
7	I5361-03DUP	1	Water			6.51	09/21/2017	16:52
8	I5365-02	1	Water			7.27	09/21/2017	17:00
9	I5368-02	1	Water			7.68	09/21/2017	17:01
10	I5369-02	1	Water			7.80	09/21/2017	17:03
11	CCV2	1	Water			12.01	09/21/2017	17:05

WORKLIST(Hardcopy Internal Chain)

2890237

WorkList Name : CORROSIONITY I5369

WorkList ID : 103139

Date : 9/21/2017 4:32:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5361-03 -1A	Corrosivity	Cool 4 deg C	PSEG02	H52	DRUM-A-B-COMP	09/21/2017	9040C
	Water	I5365-02 -1A	Corrosivity	Cool 4 deg C	PSEG01	H52	GROUND-WATER	09/21/2017	9040C
	Water	I5368-02 -1A	Corrosivity	Cool 4 deg C	PSEG01		F1123	09/21/2017	9040C
	Water	I5369-02 -1A	Corrosivity	Cool 4 deg C	PSEG01	H52	N4-1824	09/21/2017	9040C

Date/Time 09/21/17 4:32 PM

Received by: JF

Relinquished by: CR

Date/Time 09/21/17 6:25 PM

Received by: CR

Relinquished by: JF