



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90238

Temp.: 21.2 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/21/2017	16:45
2	CAL2	1	Water			7.00	09/21/2017	16:46
3	CAL3	1	Water			10.04	09/21/2017	16:47
4	ICV	1	Water			6.99	09/21/2017	16:48
5	CCV1	1	Water			2.02	09/21/2017	16:50
6	I5361-02	1	Water			6.50	09/21/2017	16:51
7	I5361-02DUP	1	Water			6.51	09/21/2017	16:52
8	I5365-01	1	Water			7.27	09/21/2017	17:00
9	I5366-01	1	Water			7.00	09/21/2017	17:15
10	I5366-02	1	Water			6.47	09/21/2017	17:16
11	I5366-03	1	Water			6.26	09/21/2017	17:18
12	I5366-04	1	Water			7.71	09/21/2017	17:19
13	I5366-05	1	Water			7.90	09/21/2017	17:20
14	I5368-01	1	Water			7.68	09/21/2017	17:01
15	I5369-01	1	Water			7.80	09/21/2017	17:03
16	CCV2	1	Water			12.01	09/21/2017	17:22

2390238

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5369

WorkList ID : 103140

Date : 9/21/2017 4:33:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5361-02 -1A	pH	Cool 4 deg C	PSEG02	H52	DRUM-A-B-COMP	09/21/2017	SM 4500-PH B
	Water	I5365-01 -1A	pH	Cool 4 deg C	PSEG01	H52	GROUND-WATER	09/21/2017	SM 4500-PH B
	Water	I5366-01 -1A	pH	Cool 4 deg C	STER07	H42	CELL-A	09/19/2017	SM 4500-PH B
	Water	I5366-02 -1A	pH	Cool 4 deg C	STER07	H42	CELLS-B-C	09/19/2017	SM 4500-PH B
	Water	I5366-03 -1A	pH	Cool 4 deg C	STER07	H42	CELL-D	09/19/2017	SM 4500-PH B
	Water	I5366-04 -1A	pH	Cool 4 deg C	STER07	H42	FW1-SUPPLY	09/19/2017	SM 4500-PH B
	Water	I5366-05 -1A	pH	Cool 4 deg C	STER07	H42	FW2-3-SUPPLY	09/19/2017	SM 4500-PH B
	Water	I5368-01 -1A	pH	Cool 4 deg C	PSEG01		F1123	09/21/2017	SM 4500-PH B
	Water	I5369-01 -1A	pH	Cool 4 deg C	PSEG01	H52	N4-1824	09/21/2017	SM 4500-PH B

Date/Time 09/21/17 4:32 PM
 Received by: JP
 Relinquished by: CP

Date/Time 09/21/17 6:30 PM
 Received by: CP
 Relinquished by: JP