


ISM02.4

LB 90252
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45647



A= 0.0000e+000

B= 2.1986e-004

C= -2.9743e-002

Rho= 0.9999830

Accept=Accepted

Slope
y=13715xcept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	-0.011	-0.011	87	0.000	87					-
0.2	0.200	0.202	0.002	1054	0.0 %	1054					1
2.5	2.500	2.535	0.035	11666	0.0 %	11666					.1
5.0	5.000	4.963	-0.037	22709	0.0 %	22709					-1
7.5	7.500	7.504	0.004	34266	0.0 %	34266					0
10.0	10.000	10.006	0.006	45647	0.0 %	45647					0

No3
9/22/12

LB90252
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	87	-		0 ISM02.4	PPB	9/22/2017 14:13	Std
	0.2 S0.2	1054	-		0.2 ISM02.4	PPB	9/22/2017 14:16	Std
	2.5 S2.5	11666	-		2.5 ISM02.4	PPB	9/22/2017 14:18	Std
	5 S5	22709	-		5 ISM02.4	PPB	9/22/2017 14:20	Std
	7.5 S7.5	34266	-		7.5 ISM02.4	PPB	9/22/2017 14:23	Std
	10 S10	45647	-		10 ISM02.4	PPB	9/22/2017 14:25	Std
ICV	ICV	17684	3.8583 -		ISM02.4	PPB	9/22/2017 14:28	SMPL
ICB	ICB	-116	-0.0552 -		ISM02.4	PPB	9/22/2017 14:30	SMPL
CCV089	CCV089	23028	5.0332 -		ISM02.4	PPB	9/22/2017 14:32	SMPL
CCB089	CCB089	-145	-0.0616 -		ISM02.4	PPB	9/22/2017 14:34	SMPL
PB102525BL	PBS004	-27	-0.0357 -		ISM02.4	PPB	9/22/2017 14:37	SMPL
I5346-01	MET1Z1	2348	0.4865 -		ISM02.4	PPB	9/22/2017 14:39	SMPL
I5346-02	MET1Z2	3519	0.744 -		ISM02.4	PPB	9/22/2017 14:41	SMPL
I5346-03	MET1Z3	3505	0.7409 -		ISM02.4	PPB	9/22/2017 14:43	SMPL
I5346-04	MET1Z4	5146	1.1017 -		ISM02.4	PPB	9/22/2017 14:46	SMPL
I5346-05	MET1Z5	4319	0.9198 -		ISM02.4	PPB	9/22/2017 14:48	SMPL
I5346-06	MET1Z6	1974	0.4043 -		ISM02.4	PPB	9/22/2017 14:50	SMPL
I5346-07	MET1Z7	1727	0.35 -		ISM02.4	PPB	9/22/2017 14:53	SMPL
I5346-08	MET1Z8	4537	0.9678 -		ISM02.4	PPB	9/22/2017 14:55	SMPL
I5346-09	MET1Z9	2226	0.4597 -		ISM02.4	PPB	9/22/2017 14:57	SMPL
I5346-10	MET200	2609	0.5439 -		ISM02.4	PPB	9/22/2017 14:59	SMPL
I5346-11	MET200D	2604	0.5428 -		ISM02.4	PPB	9/22/2017 15:02	SMPL
I5346-12	MET200S	13907	3.0279 -		ISM02.4	PPB	9/22/2017 15:04	SMPL
I5346-13	MET201	2255	0.466 -		ISM02.4	PPB	9/22/2017 15:06	SMPL
I5346-14	MET202	1832	0.373 -		ISM02.4	PPB	9/22/2017 15:09	SMPL
I5346-15	MET203	2467	0.5127 -		ISM02.4	PPB	9/22/2017 15:11	SMPL
I5346-16	MET204	6009	1.2914 -		ISM02.4	PPB	9/22/2017 15:13	SMPL
I5346-17	MET205	22020	4.8116 -		ISM02.4	PPB	9/22/2017 15:15	SMPL
I5346-18	MET206	2569	0.5351 -		ISM02.4	PPB	9/22/2017 15:18	SMPL
I5346-19	MET207	2554	0.5318 -		ISM02.4	PPB	9/22/2017 15:20	SMPL
CCV090	CCV090	22653	4.9508 -		ISM02.4	PPB	9/22/2017 15:22	SMPL
CCB090	CCB090	-139	-0.0603 -		ISM02.4	PPB	9/22/2017 15:25	SMPL
I5346-20	MET208	4640	0.9904 -		ISM02.4	PPB	9/22/2017 15:27	SMPL
I5346-21	MET2A9	1882	0.384 -		ISM02.4	PPB	9/22/2017 15:29	SMPL
I5346-22	MET2B0	4304	0.9165 -		ISM02.4	PPB	9/22/2017 15:31	SMPL
PB102526BL	PBS004	-25	-0.0352 -		ISM02.4	PPB	9/22/2017 15:34	SMPL
I5347-01	MET2B1	2391	0.4959 -		ISM02.4	PPB	9/22/2017 15:36	SMPL
I5347-02	MET2B2	3006	0.6312 -		ISM02.4	PPB	9/22/2017 15:38	SMPL
I5347-03	MET2B3	3492	0.738 -		ISM02.4	PPB	9/22/2017 15:40	SMPL
I5347-04	MET2B4	16354	3.5659 -		ISM02.4	PPB	9/22/2017 15:48	SMPL
I5347-05	MET2B4D	18399	4.0155 -		ISM02.4	PPB	9/22/2017 15:50	SMPL
I5347-06	MET2B4S	28493	6.2348 -		ISM02.4	PPB	9/22/2017 15:56	SMPL
I5347-07	MET2B5	3862	0.8194 -		ISM02.4	PPB	9/22/2017 15:58	SMPL
CCV091	CCV091	22789	4.9807 -		ISM02.4	PPB	9/22/2017 16:00	SMPL
CCB091	CCB091	-95	-0.0506 -		ISM02.4	PPB	9/22/2017 16:03	SMPL