



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB90260

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	09/22/2017	09:35
2	I5333-02		1	>100.0	>212.0	09/22/2017	10:02
3	I5333-02DUP		1	>100.0	>212.0	09/22/2017	10:30
4	I5339-02		1	>100.0	>212.0	09/22/2017	10:50
5	I5361-03		1	>100.0	>212.0	09/22/2017	11:15
6	I5365-02		1	>100.0	>212.0	09/22/2017	11:40
7	I5368-02		1	>100.0	>212.0	09/22/2017	12:05
8	I5369-02		1	>100.0	>212.0	09/22/2017	12:32

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP92217

WorkList ID : 103165

Date : 9/22/2017 8:50:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5333-02 C	Flash Point	Cool 4 deg C	PSEG01	H52	SV-30745L	09/20/2017	1010A
	Water	I5339-02 C	Flash Point	Cool 4 deg C	PSEG01	H51	FRAC-TANK-573094	09/20/2017	1010A
	Water	I5361-03 C	Flash Point	Cool 4 deg C	PSEG02	H52	DRUM-A-B-COMP	09/21/2017	1010A
	Water	I5365-02 C	Flash Point	Cool 4 deg C	PSEG01	H52	GROUND-WATER	09/21/2017	1010A
	Water	I5368-02 C	Flash Point	Cool 4 deg C	PSEG01		F1123	09/21/2017	1010A
	Water	I5369-02 C	Flash Point	Cool 4 deg C	PSEG01	H52	N4-1824	09/21/2017	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: