


7471A

LB 90261
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41242



A= 0.0000e+000

B= 2.4259e-004

C= 1.5789e-002

Rho= 0.9998551

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-50	0.000	-50				
0.2	0.200	0.186	-0.014	701	0.0 %	701				
2.5	2.500	2.590	0.090	10613	0.0 %	10613				
5.0	5.000	4.880	-0.120	20050	0.0 %	20050				
7.5	7.500	7.520	0.020	30933	0.0 %	30933	0			
10.0	10.000	10.021	0.021	41242	0.0 %	41242				

LB90261 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 22 Sep 2017 16:13:34

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.00007471A	PPB	22 Sep 2017 16:14:47	S	Std
0.2 - 1	50.2	701	-	0.20007471A	PPB	22 Sep 2017 16:17:05	S	Std
2.5 - 1	52.5	10613	-	2.50007471A	PPB	22 Sep 2017 16:19:22	S	Std
5.0 - 1	55	20050	-	5.00007471A	PPB	22 Sep 2017 16:21:40	S	Std
7.5 - 1	57.5	30933	-	7.50007471A	PPB	22 Sep 2017 16:23:57	S	Std
10.0 - 1	10	41242	-	10.00007471A	PPB	22 Sep 2017 16:28:52	S	Std
ICV70 - 1	ICV70	16678	4.0617	-7471A	PPB	22 Sep 2017 16:31:31	U	SMPL
ICB70 - 1	ICB70	-236	-0.0415	-7471A	PPB	22 Sep 2017 16:33:46	U	SMPL
CCV92 - 1	CCV92	21607	5.2574	-7471A	PPB	22 Sep 2017 16:36:02	U	SMPL
CCB92 - 1	CCB92	-225	-0.0388	-7471A	PPB	22 Sep 2017 16:38:18	U	SMPL
CRA - 1	CRA	689	0.1829	-7471A	PPB	22 Sep 2017 16:40:35	U	SMPL
HighStd - 1	HighStd	41465	10.0747	-7471A	PPB	22 Sep 2017 16:42:51	U	SMPL
ChkStd - 1	ChkStd	28468	6.9218	-7471A	PPB	22 Sep 2017 16:45:06	U	SMPL
PB102527BL - 1	PBS	-318	-0.0614	-7471A	PPB	22 Sep 2017 16:47:24	U	SMPL
PB102527BS - 1	LCS	14634	3.5658	-7471A	PPB	22 Sep 2017 16:49:43	U	SMPL
I5340-01 - 1	LT-TS-001	2893	0.7176	-7471A	PPB	22 Sep 2017 16:52:00	U	SMPL
I5340-02 - 1	LT-TS-002	3690	0.9109	-7471A	PPB	22 Sep 2017 16:54:17	U	SMPL
I5340-03 - 1	LT-TS-003	4159	1.0247	-7471A	PPB	22 Sep 2017 16:56:35	U	SMPL
I5340-04 - 1	LT-TS-004	4074	1.0041	-7471A	PPB	22 Sep 2017 16:58:52	U	SMPL
I5340-05 - 1	LT-TS-005	5593	1.3726	-7471A	PPB	22 Sep 2017 17:01:09	U	SMPL
CCV93 - 1	CCV93	20780	5.0568	-7471A	PPB	22 Sep 2017 17:03:25	U	SMPL
CCB93 - 1	CCB93	-207	-0.0344	-7471A	PPB	22 Sep 2017 17:05:41	U	SMPL
I5340-05DUP - 1	LT-TS-005DUP	4937	1.2134	-7471A	PPB	22 Sep 2017 17:07:57	U	SMPL
I5340-05MS - 1	LT-TS-005MS	20502	4.9893	-7471A	PPB	22 Sep 2017 17:10:13	U	SMPL
I5340-05MSD - 1	LT-TS-005MSD	20916	5.0898	-7471A	PPB	22 Sep 2017 17:12:29	U	SMPL
I5340-06 - 1	LT-TS-006	2923	0.7249	-7471A	PPB	22 Sep 2017 17:14:45	U	SMPL
I5340-07 - 1	LT-TS-007	3461	0.8554	-7471A	PPB	22 Sep 2017 17:17:03	U	SMPL
I5340-08 - 1	LT-TS-008	2989	0.7409	-7471A	PPB	22 Sep 2017 17:19:20	U	SMPL
I5340-09 - 1	LT-TS-009	3821	0.9427	-7471A	PPB	22 Sep 2017 17:21:37	U	SMPL
I5340-10 - 1	LT-TS-010	3012	0.7465	-7471A	PPB	22 Sep 2017 17:23:55	U	SMPL
I5340-11 - 1	LT-TS-011	3711	0.9160	-7471A	PPB	22 Sep 2017 17:26:12	U	SMPL
I5340-12 - 1	LT-TS-012	9626	2.3509	-7471A	PPB	22 Sep 2017 17:28:28	U	SMPL
CCV94 - 1	CCV94	20407	4.9663	-7471A	PPB	22 Sep 2017 17:30:45	U	SMPL
CCB94 - 1	CCB94	-366	-0.0730	-7471A	PPB	22 Sep 2017 17:33:01	U	SMPL
I5364-01 - 1	SASP2P-ENV-118-5(0-5)	3613	0.8923	-7471A	PPB	22 Sep 2017 17:35:20	U	SMPL
I5364-02 - 1	SASP2P-ENV-118-5(5-10)	7996	1.9555	-7471A	PPB	22 Sep 2017 17:37:36	U	SMPL
I5364-03 - 1	SASP2P-ENV-125-17(0-5)	8868	2.1671	-7471A	PPB	22 Sep 2017 17:39:52	U	SMPL
I5364-04 - 1	SASP2P-ENV-125-17(5-10)	577	0.1558	-7471A	PPB	22 Sep 2017 17:42:09	U	SMPL
I5364-05 - 1	SASP2P-ENV-116-6(0-5)	4843	1.1906	-7471A	PPB	22 Sep 2017 17:44:25	U	SMPL
I5364-06 - 1	SASP2P-ENV-116-6(5-10)	7094	1.7367	-7471A	PPB	22 Sep 2017 17:46:42	U	SMPL
I5340-05LX5 - 1		941	0.2441	-7471A	PPB	22 Sep 2017 17:48:59	U	SMPL
I5340-05A - 1		21982	5.3484	-7471A	PPB	22 Sep 2017 17:51:17	U	SMPL
CCV95 - 1	CCV95	20657	5.0269	-7471A	PPB	22 Sep 2017 17:53:35	U	SMPL
CCB95 - 1	CCB95	-241	-0.0427	-7471A	PPB	22 Sep 2017 17:55:51	U	SMPL