



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90262

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP58201
TURBIDITY CCV 10 NTU	WP58202
Turbidity std, 40NTU	WP58196
turbidity std 5 NTU	WP58199
turbidity std 20 NTU	WP58197
Turbidity std 1 NTU	WP58200
turbidity std 0 NTU	WP58203
turbidity std 10 NTU	WP58198

Intercept: 0.0091

Slope: 0.9927

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.080	09/22/2017	15:20
2	CAL2	1	1	1.030	09/22/2017	15:23
3	CAL3	5	1	4.950	09/22/2017	15:26
4	CAL4	10	1	9.940	09/22/2017	15:28
5	CAL5	20	1	19.700	09/22/2017	15:31
6	CAL6	40	1	39.800	09/22/2017	15:33



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90262

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.600	9.661	09/22/2017	15:37
2	ICB		1	0.079	0.070	09/22/2017	15:40
3	CCV1	10	1	9.850	9.913	09/22/2017	15:43
4	CCB1		1	0.080	0.071	09/22/2017	15:46
5	LB90262BL		1	0.080	0.071	09/22/2017	15:49
6	I5381-01		1	35.700	35.953	09/22/2017	15:51
7	I5381-01DUP		1	35.600	35.853	09/22/2017	15:54
8	I5381-02		1	11.800	11.878	09/22/2017	15:58
9	I5381-03		1	6.230	6.267	09/22/2017	16:01
10	I5381-04		1	36.400	36.659	09/22/2017	16:03
11	I5381-05		1	6.010	6.045	09/22/2017	16:06
12	I5381-06		1	6.390	6.428	09/22/2017	16:09
13	I5381-07		1	12.400	12.482	09/22/2017	16:12
14	I5381-08		1	15.400	15.504	09/22/2017	16:15
15	CCV2	10	1	9.750	9.813	09/22/2017	16:17
16	CCB2		1	0.081	0.072	09/22/2017	16:19
17	I5381-09		1	14.900	15.000	09/22/2017	16:22
18	I5381-10		1	5.300	5.330	09/22/2017	16:25
19	I5381-11		1	4.820	4.846	09/22/2017	16:28
20	I5381-12		1	5.180	5.209	09/22/2017	16:31
21	I5387-07		1	0.192	0.184	09/22/2017	16:33
22	I5387-08		1	0.197	0.189	09/22/2017	16:36
23	I5387-09		1	0.177	0.169	09/22/2017	16:38
24	I5387-10		1	0.220	0.212	09/22/2017	16:41
25	I5387-11		1	0.214	0.206	09/22/2017	16:43
26	I5387-12		1	0.177	0.169	09/22/2017	16:46
27	CCV3	10	1	9.540	9.601	09/22/2017	16:49
28	CCB3		1	0.082	0.073	09/22/2017	16:51

LB 90262

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR092217

WorkList ID : 103178

Date : 9/22/2017 12:24:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5381-01	Turbidity	Cool 4 deg C	HDR103	H32	20170921-092006	09/21/2017	SM2130 B
	Water	I5381-02	Turbidity	Cool 4 deg C	HDR103	H32	20170921-092529	09/21/2017	SM2130 B
	Water	I5381-03	Turbidity	Cool 4 deg C	HDR103	H32	20170921-092732	09/21/2017	SM2130 B
	Water	I5381-04	Turbidity	Cool 4 deg C	HDR103	H32	20170921-102649	09/21/2017	SM2130 B
	Water	I5381-05	Turbidity	Cool 4 deg C	HDR103	H32	20170921-103110	09/21/2017	SM2130 B
	Water	I5381-06	Turbidity	Cool 4 deg C	HDR103	H32	20170921-103319	09/21/2017	SM2130 B
	Water	I5381-07	Turbidity	Cool 4 deg C	HDR103	H32	20170921-135234	09/21/2017	SM2130 B
	Water	I5381-08	Turbidity	Cool 4 deg C	HDR103	H32	20170921-135654	09/21/2017	SM2130 B
	Water	I5381-09	Turbidity	Cool 4 deg C	HDR103	H32	20170921-135850	09/21/2017	SM2130 B
	Water	I5381-10	Turbidity	Cool 4 deg C	HDR103	H32	20170921-150548	09/21/2017	SM2130 B
	Water	I5381-11	Turbidity	Cool 4 deg C	HDR103	H32	20170921-151059	09/21/2017	SM2130 B
	Water	I5381-12	Turbidity	Cool 4 deg C	HDR103	H32	20170921-151227	09/21/2017	SM2130 B
	Water	I5387-07	Turbidity	Cool 4 deg C	HILL02	H13	AC01A2	09/22/2017	SM2130 B
	Water	I5387-08	Turbidity	Cool 4 deg C	HILL02	H13	AC01B2	09/22/2017	SM2130 B
	Water	I5387-09	Turbidity	Cool 4 deg C	HILL02	H13	AC02A2	09/22/2017	SM2130 B
	Water	I5387-10	Turbidity	Cool 4 deg C	HILL02	H13	AC02B2	09/22/2017	SM2130 B
	Water	I5387-11	Turbidity	Cool 4 deg C	HILL02	H13	AC03A2	09/22/2017	SM2130 B
	Water	I5387-12	Turbidity	Cool 4 deg C	HILL02	H13	AC03B2	09/22/2017	SM2130 B

Date/Time 9/22/17 16:20

Received by: AK

Relinquished by: CR

Date/Time

Received by:

Relinquished by: AK