



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90266

Temp.: 19.2 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/22/2017	16:00
2	CAL2	1	Water			7.01	09/22/2017	16:01
3	CAL3	1	Water			10.03	09/22/2017	16:03
4	ICV	1	Water			7.00	09/22/2017	16:05
5	CCV1	1	Water			2.02	09/22/2017	16:06
6	I5383-01	1	Water			4.48	09/22/2017	16:08
7	I5383-01DUP	1	Water			4.49	09/22/2017	16:10
8	CCV2	1	Water			12.01	09/22/2017	16:11

2B90266

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5337

WorkList ID : 103183

Date : 9/22/2017 3:41:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5383-01-A	pH	Cool 4 deg C	STER12	H21	CELL-POOL-WATER	09/21/2017	SM 4500-PH B

Date/Time 09/22/17 3:41 PM
 Received by: JA CR
 Relinquished by:

Date/Time 09/22/17 5:30 PM
 Received by: CR
 Relinquished by: JA