



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90267

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/22/2017	16:15
2	CAL2	1	Water			7.01	09/22/2017	16:17
3	CAL3	1	Water			10.03	09/22/2017	16:19
4	ICV	1	Water			6.99	09/22/2017	16:20
5	CCV1	1	Water			2.01	09/22/2017	16:21
6	I5388-01	1	Solid	20.02	20	5.92	09/22/2017	16:22
7	I5388-01DUP	1	Solid	20.03	20	5.93	09/22/2017	16:25
8	I5391-01	1	Solid	20.01	20	6.09	09/22/2017	16:26
9	CCV2	1	Water			12.01	09/22/2017	16:27

2B9026F

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5888

WorkList ID : 103182

Date : 9/22/2017 3:40:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5388-01-17	pH	Cool 4 deg C	TULL02	H61	FK-PS-01-016	09/22/2017	9045C
	Solid	I5391-01 - 17	pH	Cool 4 deg C	TULL02	H11	FK-PS-01-015	09/22/2017	9045C

Date/Time 09/22/17 3:41 PM
Received by: JP
Relinquished by: JP

Date/Time 09/22/17 4:30 PM
Received by: CP
Relinquished by: JP