

245.1


LB 90271

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32035



A= 0.0000e+000

B= 3.1060e-004

C= 8.8322e-002

Rho= 0.9998720

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	-120	0.000	-120				
0.2	0.200	0.240	0.040	487	0.0 %	487				
2.5	2.500	2.440	-0.060	7573	0.0 %	7573				
5.0	5.000	4.897	-0.103	15482	0.0 %	15482				
7.5	7.500	7.533	0.033	23970	0.0 %	23970				
10.0	10.000	10.038	0.038	32035	0.0 %	32035				

LB90271 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Sep 2017 18:08:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-120	-	0.0000	245.1	PPB	22 Sep 2017 18:48:51	S	Std
0.2 - 1	50.2	487	-	0.2000	245.1	PPB	22 Sep 2017 18:50:59	S	Std
2.5 - 1	52.5	7573	-	2.5000	245.1	PPB	22 Sep 2017 18:53:08	S	Std
5.0 - 1	55	15482	-	5.0000	245.1	PPB	22 Sep 2017 18:55:16	S	Std
7.5 - 1	57.5	23970	-	7.5000	245.1	PPB	22 Sep 2017 18:57:24	S	Std
10.0 - 1	510	32035	-	10.0000	245.1	PPB	22 Sep 2017 18:59:33	S	Std
ICV71 - 1	ICV71	12695	4.0314	-245.1		PPB	22 Sep 2017 19:02:14	U	SMPL
ICB71 - 1	ICB71	53	0.1048	-245.1		PPB	22 Sep 2017 19:04:21	U	SMPL
CCV96 - 1	CCV96	15759	4.9831	-245.1		PPB	22 Sep 2017 19:06:27	U	SMPL
CCB96 - 1	CCB96	-166	0.0368	-245.1		PPB	22 Sep 2017 19:08:34	U	SMPL
CRA - 1	CRA	458	0.2306	-245.1		PPB	22 Sep 2017 19:10:41	U	SMPL
HighStd - 1	HighStd	31765	9.9545	-245.1		PPB	22 Sep 2017 19:12:48	U	SMPL
ChkStd - 1	ChkStd	22199	6.9833	-245.1		PPB	22 Sep 2017 19:14:55	U	SMPL
PB102549BL - 1	PBW	-254	0.0094	-245.1		PPB	22 Sep 2017 19:17:03	U	SMPL
PB102549BS - 1	LCSW	12735	4.0438	-245.1		PPB	22 Sep 2017 19:21:38	U	SMPL
I5307-01 - 1	EFF-DAY-1-METAL	1113	0.4340	-245.1		PPB	22 Sep 2017 19:23:45	U	SMPL
I5307-01DUP - 1	EFF-DAY-1-METALDUP	975	0.3912	-245.1		PPB	22 Sep 2017 19:25:54	U	SMPL
I5307-01MS - 1	EFF-DAY-1-METALMS	13623	4.3196	-245.1		PPB	22 Sep 2017 19:28:03	U	SMPL
I5307-01MSD - 1	EFF-DAY-1-METALMSD	13067	4.1469	-245.1		PPB	22 Sep 2017 19:30:11	U	SMPL
I5387-01 - 1	AC01A1	269	0.1719	-245.1		PPB	22 Sep 2017 19:32:18	U	SMPL
CCV97 - 1	CCV97	15260	4.8281	-245.1		PPB	22 Sep 2017 19:34:26	U	SMPL
CCB97 - 1	CCB97	-179	0.0327	-245.1		PPB	22 Sep 2017 19:36:33	U	SMPL
I5387-02 - 1	AC01B1	538	0.2554	-245.1		PPB	22 Sep 2017 19:38:40	U	SMPL
I5387-03 - 1	AC02A1	950	0.3834	-245.1		PPB	22 Sep 2017 19:40:47	U	SMPL
I5387-04 - 1	AC02B1	1164	0.4499	-245.1		PPB	22 Sep 2017 19:42:54	U	SMPL
I5387-05 - 1	AC03A1	1333	0.5024	-245.1		PPB	22 Sep 2017 19:45:01	U	SMPL
I5387-06 - 1	AC03B1	1619	0.5912	-245.1		PPB	22 Sep 2017 19:47:08	U	SMPL
I5307-01LX5 - 1		152	0.1355	-245.1		PPB	22 Sep 2017 19:52:06	U	SMPL
I5307-01A - 1		15069	4.7688	-245.1		PPB	22 Sep 2017 19:54:13	U	SMPL
CCV98 - 1	CCV98	15314	4.8449	-245.1		PPB	22 Sep 2017 19:56:21	U	SMPL
CCB98 - 1	CCB98	-29	0.0793	-245.1		PPB	22 Sep 2017 19:58:30	U	SMPL