



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90272

Reagent/Standard	Lot/Log #
TURBIDITY ICV 10 NTU	WP58201
TURBIDITY CCV 10 NTU	WP58202
Turbidity std, 40NTU	WP58196
turbidity std 5 NTU	WP58199
Turbidity std 1 NTU	WP58200
turbidity std 0 NTU	WP58203
turbidity std 10 NTU	WP58198

Intercept: 0.0091

Slope: 0.9927

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.080	09/22/2017	15:20
2	CAL2	1	1	1.030	09/22/2017	15:23
3	CAL3	5	1	4.950	09/22/2017	15:26
4	CAL4	10	1	9.940	09/22/2017	15:28
5	CAL5	20	1	19.700	09/22/2017	15:31
6	CAL6	40	1	39.800	09/22/2017	15:33



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB90272

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.600	9.661	09/22/2017	15:37
2	ICB		1	0.070	0.061	09/22/2017	15:40
3	CCV1	10	1	9.850	9.913	09/22/2017	15:43
4	CCB1		1	0.080	0.071	09/22/2017	15:46
5	LB90272BL		1	0.080	0.071	09/22/2017	15:49
6	I5397-01		1	16.200	16.310	09/22/2017	17:06
7	I5397-01DUP		1	16.000	16.108	09/22/2017	17:09
8	I5397-02		1	30.400	30.614	09/22/2017	17:12
9	I5397-03		1	6.180	6.216	09/22/2017	17:15
10	I5397-04		1	23.400	23.563	09/22/2017	17:17
11	I5397-05		1	25.800	25.981	09/22/2017	17:19
12	I5397-06		1	8.940	8.997	09/22/2017	17:21
13	I5397-07		1	31.300	31.521	09/22/2017	17:23
14	I5397-08		1	36.200	36.457	09/22/2017	17:25
15	CCV2	10	1	9.810	9.873	09/22/2017	17:27
16	CCB2		1	0.082	0.073	09/22/2017	17:29
17	I5397-09		1	38.500	38.774	09/22/2017	17:31
18	I5397-10		1	38.300	38.572	09/22/2017	17:33
19	I5397-11		1	21.700	21.850	09/22/2017	17:35
20	I5397-12		1	24.100	24.268	09/22/2017	17:37
21	I5397-13		1	23.500	23.664	09/22/2017	17:40
22	I5397-14		1	19.600	19.735	09/22/2017	17:42
23	I5397-15		1	26.500	26.686	09/22/2017	17:45
24	I5397-16		1	24.100	24.268	09/22/2017	17:48
25	I5397-17		1	29.400	29.607	09/22/2017	17:50
26	I5397-18		1	31.500	31.722	09/22/2017	17:52
27	CCV3	10	1	9.800	9.863	09/22/2017	17:53
28	CCB3		1	0.079	0.070	09/22/2017	17:54
29	I5397-19		1	31.800	32.025	09/22/2017	17:56
30	I5397-20		1	23.600	23.764	09/22/2017	17:58
31	CCV4	10	1	9.760	9.823	09/22/2017	18:00
32	CCB4		1	0.081	0.072	09/22/2017	18:03

6890242

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR092217-2

WorkList ID : 103186

Date : 9/22/2017 4:23:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5397-01	Turbidity	Cool 4 deg C	HDR103	H22	20170922-091918	09/22/2017	SM2130 B
	Water	I5397-02	Turbidity	Cool 4 deg C	HDR103	H22	20170922-092451	09/22/2017	SM2130 B
	Water	I5397-03	Turbidity	Cool 4 deg C	HDR103	H22	20170922-092946	09/22/2017	SM2130 B
	Water	I5397-04	Turbidity	Cool 4 deg C	HDR103	H22	20170922-120119	09/22/2017	SM2130 B
	Water	I5397-05	Turbidity	Cool 4 deg C	HDR103	H22	20170922-120416	09/22/2017	SM2130 B
	Water	I5397-06	Turbidity	Cool 4 deg C	HDR103	H22	20170922-120657	09/22/2017	SM2130 B
	Water	I5397-19	Turbidity	Cool 4 deg C	HDR103	H22	20170922-114323	09/22/2017	SM2130 B
	Water	I5397-20	Turbidity	Cool 4 deg C	HDR103	H22	20170922-114332	09/22/2017	SM2130 B
	Water	I5397-13	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113427	09/22/2017	SM2130 B
	Water	I5397-14	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113432	09/22/2017	SM2130 B
	Water	I5397-15	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113444	09/22/2017	SM2130 B
	Water	I5397-16	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113450	09/22/2017	SM2130 B
	Water	I5397-17	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113455	09/22/2017	SM2130 B
	Water	I5397-18	Turbidity	Cool 4 deg C	HDR103	H22	20170922-114319	09/22/2017	SM2130 B
	Water	I5397-07	Turbidity	Cool 4 deg C	HDR103	H22	20170922-112332	09/22/2017	SM2130 B
	Water	I5397-08	Turbidity	Cool 4 deg C	HDR103	H22	20170922-112348	09/22/2017	SM2130 B
	Water	I5397-09	Turbidity	Cool 4 deg C	HDR103	H22	20170922-112400	09/22/2017	SM2130 B
	Water	I5397-10	Turbidity	Cool 4 deg C	HDR103	H22	20170922-112405	09/22/2017	SM2130 B
	Water	I5397-11	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113357	09/22/2017	SM2130 B
	Water	I5397-12	Turbidity	Cool 4 deg C	HDR103	H22	20170922-113402	09/22/2017	SM2130 B

9/22/17 17:00

Date/Time

AK

Received by:

GA

Relinquished by:

9/22/17 18:05

Date/Time

Received by:

AK

Relinquished by: