



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90274

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/22/2017	17:30
2	CAL2	1	Water			7.00	09/22/2017	17:31
3	CAL3	1	Water			10.03	09/22/2017	17:33
4	ICV	1	Water			7.01	09/22/2017	17:35
5	CCV1	1	Water			2.02	09/22/2017	17:36
6	I5380-03	1	Solid	20.02	20	5.89	09/22/2017	17:37
7	I5394-03	1	Solid	20.03	20	10.63	09/22/2017	17:38
8	I5394-06	1	Solid	20.01	20	8.71	09/22/2017	17:39
9	I5394-09	1	Solid	20.04	20	7.75	09/22/2017	17:40
10	I5395-03	1	Solid	20.02	20	5.48	09/22/2017	17:41
11	I5395-06	1	Solid	20.03	20	5.57	09/22/2017	17:42
12	I5395-09	1	Solid	20.01	20	5.83	09/22/2017	17:44
13	I5395-12	1	Solid	20.04	20	6.05	09/22/2017	17:45
14	I5395-15	1	Solid	20.02	20	6.07	09/22/2017	17:46
15	I5395-18	1	Solid	20.03	20	7.73	09/22/2017	17:48
16	CCV2	1	Water			12.01	09/22/2017	17:50
17	I5395-18DUP	1	Solid	20.02	20	7.74	09/22/2017	17:51
18	CCV3	1	Water			12.02	09/22/2017	17:55

1290274

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5380 WorkList ID : 103189 Date : 9/22/2017 5:23:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5380-03	Corrosivity	Cool 4 deg C	PSEG05	H52	SU-092217-04-COMP	09/22/2017	9045C
	Solid	I5394-03	Corrosivity	Cool 4 deg C	PSEG01	H52	15-50	09/22/2017	9045C
	Solid	I5394-06	Corrosivity	Cool 4 deg C	PSEG01	H52	20-41	09/22/2017	9045C
	Solid	I5394-09	Corrosivity	Cool 4 deg C	PSEG01	H52	RT-3992	09/22/2017	9045C
	Solid	I5395-03	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-5-0-5	09/22/2017	9045C
	Solid	I5395-06	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-4-0-5	09/22/2017	9045C
	Solid	I5395-09	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-3-0-5	09/22/2017	9045C
	Solid	I5395-12	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-2-0-5	09/22/2017	9045C
	Solid	I5395-15	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-1-0-5	09/22/2017	9045C
	Solid	I5395-18	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-6-0-5	09/22/2017	9045C

Date/Time 09/22/17 5:00 PM
 Received by: JP
 Relinquished by: JP

Date/Time 09/22/17 7:55 PM
 Received by: JP
 Relinquished by: JP