



## Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90290

Temp.: 21.8 °C

SLOPE.: 98.2

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2183         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2154         |
| buffer solution pH 7 yellow    | W2219         |
| pH 2.00 Buffer                 | W2079         |
| Buffer Solution, PH12 (500ml)  | W2267         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight<br>(gm) | Volume<br>(ml) | Result<br>(pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|----------------|----------------|----------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |                |                | 4.01           | 09/25/2017 | 11:55     |
| 2   | CAL2        | 1  | Water  |                |                | 7.01           | 09/25/2017 | 11:56     |
| 3   | CAL3        | 1  | Water  |                |                | 10.03          | 09/25/2017 | 12:00     |
| 4   | ICV         | 1  | Water  |                |                | 7.01           | 09/25/2017 | 12:02     |
| 5   | CCV1        | 1  | Water  |                |                | 2.01           | 09/25/2017 | 12:05     |
| 6   | I5413-01    | 1  | Water  |                |                | 4.26           | 09/25/2017 | 12:06     |
| 7   | I5413-01DUP | 1  | Water  |                |                | 4.27           | 09/25/2017 | 12:07     |
| 8   | CCV2        | 1  | Water  |                |                | 12.01          | 09/25/2017 | 12:10     |

LB 90240

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5413

WorkList ID : 103245

Date : 9/25/2017 11:47:35 AM

| Due Date | Matrix | Sample     | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|----------|--------|------------|------|--------------|----------|------------------|-----------------|--------------|--------------|
|          | Water  | I5413-01-A | pH   | NONE         | STER08   | H61              | POOLWATER       | 09/07/2017   | SM 4500-PH B |

Date/Time 09/25/17 11:47 AM  
Received by: JP  
Relinquished by: JP

Date/Time 09/25/17 12:45 PM  
Received by: CF  
Relinquished by: JP