


7470A

LB 90297
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34810



A= 0.0000e+000

B= 2.8552e-004

C= 2.8815e-002

Rho= 0.9999290

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-137	0.000	-137				
0.2	0.200	0.231	0.031	708	0.0 %	708				
2.5	2.500	2.427	-0.073	8399	0.0 %	8399				
5.0	5.000	5.059	0.059	17618	0.0 %	17618				
7.5	7.500	7.526	0.026	26257	0.0 %	26257				
10.0	10.000	9.968	-0.032	34810	0.0 %	34810				

LB90297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Sep 2017 14:41:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-137	-	0.0000	7470A	PPB	25 Sep 2017 14:56:18	S	Std
0.2 - 1	502	708	-	0.2000	7470A	PPB	25 Sep 2017 14:58:26	S	Std
2.5 - 1	52.5	8399	-	2.5000	7470A	PPB	25 Sep 2017 15:00:35	S	Std
5.0 - 1	55	17618	-	5.0000	7470A	PPB	25 Sep 2017 15:02:43	S	Std
7.5 - 1	57.5	26257	-	7.5000	7470A	PPB	25 Sep 2017 15:04:51	S	Std
10.0 - 1	510	34810	-	10.0000	7470A	PPB	25 Sep 2017 15:09:27	S	Std
ICV72 - 1	ICV72	13921	4.0035	-7470A	PPB	25 Sep 2017 15:11:56	U	SMPL	
ICB72 - 1	ICB72	-175	-0.0212	-7470A	PPB	25 Sep 2017 15:14:02	U	SMPL	
CCV01 - 1	CCV01	16907	4.8561	-7470A	PPB	25 Sep 2017 15:16:08	U	SMPL	
CCB01 - 1	CCB01	-149	-0.0137	-7470A	PPB	25 Sep 2017 15:18:14	U	SMPL	
CRA - 1	CRA	618	0.2053	-7470A	PPB	25 Sep 2017 15:20:21	U	SMPL	
HighStd - 1	HighStd	34048	9.7501	-7470A	PPB	25 Sep 2017 15:22:28	U	SMPL	
ChkStd - 1	ChkStd	24825	7.1168	-7470A	PPB	25 Sep 2017 15:24:34	U	SMPL	
PB102547BL - 1	PBW	-247	-0.0417	-7470A	PPB	25 Sep 2017 15:26:43	U	SMPL	
PB102547BS - 1	LCSW	13820	3.9747	-7470A	PPB	25 Sep 2017 15:28:52	U	SMPL	
PB102547TB - 1	PB102523TB	-157	-0.0160	-7470A	PPB	25 Sep 2017 15:30:59	U	SMPL	
I5361-03 - 1	DRUM-A-B-COMP	-93	0.0023	-7470A	PPB	25 Sep 2017 15:33:08	U	SMPL	
I5365-02 - 1	GROUND-WATER	-47	0.0154	-7470A	PPB	25 Sep 2017 15:35:16	U	SMPL	
I5368-02 - 1	F1123	-55	0.0131	-7470A	PPB	25 Sep 2017 15:37:25	U	SMPL	
I5369-02 - 1	N4-1824	-45	0.0160	-7470A	PPB	25 Sep 2017 15:39:31	U	SMPL	
CCV02 - 1	CCV02	16732	4.8061	-7470A	PPB	25 Sep 2017 15:41:38	U	SMPL	
CCB02 - 1	CCB02	-157	-0.0160	-7470A	PPB	25 Sep 2017 15:43:45	U	SMPL	
I5369-02DUP - 1	N4-1824DUP	-58	0.0123	-7470A	PPB	25 Sep 2017 15:45:52	U	SMPL	
I5369-02MS - 1	N4-1824MS	13392	3.8525	-7470A	PPB	25 Sep 2017 15:47:58	U	SMPL	
I5369-02MSD - 1	N4-1824MSD	12984	3.7360	-7470A	PPB	25 Sep 2017 15:50:05	U	SMPL	
PB102548BL - 1	PBW	-153	-0.0149	-7470A	PPB	25 Sep 2017 15:52:12	U	SMPL	
PB102548BS - 1	LCSW	13617	3.9167	-7470A	PPB	25 Sep 2017 15:54:19	U	SMPL	
I5361-02 - 1	DRUM-A-B-COMP	172	0.0779	-7470A	PPB	25 Sep 2017 15:56:27	U	SMPL	
I5365-01 - 1	GROUND-WATER	-206	-0.0300	-7470A	PPB	25 Sep 2017 15:58:35	U	SMPL	
I5368-01 - 1	F1123	-83	0.0051	-7470A	PPB	25 Sep 2017 16:00:43	U	SMPL	
I5369-01 - 1	N4-1824	763	0.2467	-7470A	PPB	25 Sep 2017 16:02:52	U	SMPL	
I5369-01DUP - 1	N4-1824DUP	833	0.2667	-7470A	PPB	25 Sep 2017 16:04:59	U	SMPL	
CCV03 - 1	CCV03	17695	5.0811	-7470A	PPB	25 Sep 2017 16:07:06	U	SMPL	
CCB03 - 1	CCB03	-167	-0.0189	-7470A	PPB	25 Sep 2017 16:09:13	U	SMPL	
I5369-01MS - 1	N4-1824MS	15692	4.5092	-7470A	PPB	25 Sep 2017 16:11:20	U	SMPL	
I5369-01MSD - 1	N4-1824MSD	15697	4.5106	-7470A	PPB	25 Sep 2017 16:13:27	U	SMPL	
I5369-02LX5 - 1		-139	-0.0109	-7470A	PPB	25 Sep 2017 16:15:34	U	SMPL	
I5369-02A - 1		14703	4.2268	-7470A	PPB	25 Sep 2017 16:17:41	U	SMPL	
I5369-01LX5 - 1		62	0.0465	-7470A	PPB	25 Sep 2017 16:19:48	U	SMPL	
I5369-01A - 1		16495	4.7384	-7470A	PPB	25 Sep 2017 16:21:56	U	SMPL	
CCV04 - 1	CCV04	17624	5.0608	-7470A	PPB	25 Sep 2017 16:24:04	U	SMPL	
CCB04 - 1	CCB04	-190	-0.0254	-7470A	PPB	25 Sep 2017 16:26:12	U	SMPL	