



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90306

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/25/2017	17:15
2	CAL2	1	Water			7.00	09/25/2017	17:16
3	CAL3	1	Water			10.03	09/25/2017	17:19
4	ICV	1	Water			7.00	09/25/2017	17:21
5	CCV1	1	Water			2.01	09/25/2017	17:23
6	I5415-01	1	Solid	20.02	20	5.65	09/25/2017	17:25
7	I5421-02	1	Solid	20.03	20	10.03	09/25/2017	17:26
8	I5423-03	1	Solid	20.04	20	6.83	09/25/2017	17:29
9	I5423-03DUP	1	Solid	20.03	20	6.84	09/25/2017	17:30
10	I5423-06	1	Solid	20.01	20	7.03	09/25/2017	17:32
11	I5428-02	1	Solid	20.03	20	8.06	09/25/2017	17:33
12	CCV2	1	Water			12.01	09/25/2017	17:35

WORKLIST(Hardcopy Internal Chain)

2390306

WorkList Name : CORROSIONITY I5415

WorkList ID : 103255

Date : 9/25/2017 1:12:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/28/2017	Solid	I5415-01	Corrosivity	Cool 4 deg C	PSEG01	H52	OILY-RAGS-DEBRIS	09/25/2017	9045C
09/28/2017	Solid	I5421-02	Corrosivity	Cool 4 deg C	PSEG01	H52	JC-DRUM	09/25/2017	9045C
09/28/2017	Solid	I5423-03	Corrosivity	Cool 4 deg C	PSEG05	H52	SA-06-092517-A	09/25/2017	9045C
09/28/2017	Solid	I5423-06	Corrosivity	Cool 4 deg C	PSEG05	H52	SA-06-092517-B	09/25/2017	9045C
	Solid	I5428-02	Corrosivity	Cool 4 deg C	PSEG02	H52	A-B-C-DRUM-COMP	09/25/2017	9045C

Date/Time 09/25/17 1:30 PM

Received by: JP

Relinquished by: _____

Date/Time 09/25/17 5:45 PM

Received by: _____

Relinquished by: JP