



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90307

Temp.: 23.5 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/25/2017	17:35
2	CAL2	1	Water			7.01	09/25/2017	17:36
3	CAL3	1	Water			10.03	09/25/2017	17:38
4	ICV	1	Water			6.99	09/25/2017	17:39
5	CCV1	1	Water			2.02	09/25/2017	17:40
6	I5429-02	1	Water			7.60	09/25/2017	17:41
7	I5429-02DUP	1	Water			7.61	09/25/2017	17:42
8	CCV2	1	Water			12.01	09/25/2017	17:44

WORKLIST(Hardcopy Internal Chain)

2B90307

WorkList Name : CORROSIVITY i5429

WorkList ID : 103259

Date : 9/25/2017 5:17:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2017	Water	I5429-02	Corrosivity	Cool 4 deg C	PSEG01	H52	SV-26493-LM	09/25/2017	9040C

Date/Time 09/25/17 5:20 PM
 Received by: JP
 Relinquished by: CR

Date/Time 09/25/17 6:00 PM
 Received by: CR
 Relinquished by: JP