


ISM02.4

LB 90311
INSTRUMENT ID: CV2

Linear

 μ Abs.:

40652



A= 0.0000e+000

B= 2.4601e-004

C= -8.4727e-004

Rho= 0.9999792

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	0.004	0.004	20	0.000	20					-
0.2	0.200	0.188	-0.012	766	0.0 %	766					-6
2.5	2.500	2.534	0.034	10303	0.0 %	10303					-1
5.0	5.000	4.958	-0.042	20156	0.0 %	20156					-1
7.5	7.500	7.517	0.017	30560	0.0 %	30560					0
10.0	10.000	10.000	-0.000	40652	0.0 %	40652					0

NB
9/26/12

LB90311
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	20	-		0 ISM02.4	PPB	9/26/2017 8:51	Std
	0.2 S0.2	766	-		0.2 ISM02.4	PPB	9/26/2017 8:53	Std
	2.5 S2.5	10303	-		2.5 ISM02.4	PPB	9/26/2017 8:56	Std
	5 S5	20156	-		5 ISM02.4	PPB	9/26/2017 8:58	Std
	7.5 S7.5	30560	-		7.5 ISM02.4	PPB	9/26/2017 9:00	Std
	10 S10	40652	-		10 ISM02.4	PPB	9/26/2017 9:05	Std
ICV	ICV	15797	3.8853 -		ISM02.4	PPB	9/26/2017 9:08	SMPL
ICB	ICB	-209	-0.0523 -		ISM02.4	PPB	9/26/2017 9:10	SMPL
CCV005	CCV005	20330	5.0005 -		ISM02.4	PPB	9/26/2017 9:13	SMPL
CCB005	CCB005	-134	-0.0338 -		ISM02.4	PPB	9/26/2017 9:15	SMPL
PB102597BL	PBW003	-51	-0.0134 -		ISM02.4	PPB	9/26/2017 9:17	SMPL
I5332-01	MBDA73	-35	-0.0095 -		ISM02.4	PPB	9/26/2017 9:19	SMPL
I5332-02	MBDA74	-268	-0.0668 -		ISM02.4	PPB	9/26/2017 9:22	SMPL
I5332-03	MBDA75	-210	-0.0525 -		ISM02.4	PPB	9/26/2017 9:24	SMPL
I5332-04	MBDA76	-167	-0.0419 -		ISM02.4	PPB	9/26/2017 9:26	SMPL
I5332-05	MBDA77	-128	-0.0323 -		ISM02.4	PPB	9/26/2017 9:29	SMPL
I5332-06	MBDA78	-154	-0.0387 -		ISM02.4	PPB	9/26/2017 9:31	SMPL
I5332-07	MBDA78D	-208	-0.052 -		ISM02.4	PPB	9/26/2017 9:33	SMPL
I5332-08	MBDA78S	3942	0.9689 -		ISM02.4	PPB	9/26/2017 9:35	SMPL
I5332-09	MBDA81	-248	-0.0619 -		ISM02.4	PPB	9/26/2017 9:38	SMPL
I5332-10	MBDA88	48	0.011 -		ISM02.4	PPB	9/26/2017 9:40	SMPL
I5332-11	MBDA89	1061	0.2602 -		ISM02.4	PPB	9/26/2017 9:42	SMPL
I5332-12	MBDA90	-17	-0.005 -		ISM02.4	PPB	9/26/2017 9:44	SMPL
I5332-13	MBDA91	-202	-0.0505 -		ISM02.4	PPB	9/26/2017 9:47	SMPL
I5332-14	MBDA93	-258	-0.0643 -		ISM02.4	PPB	9/26/2017 9:49	SMPL
I5332-15	MBDA94	-249	-0.0621 -		ISM02.4	PPB	9/26/2017 9:51	SMPL
I5332-16	MBDA95	-244	-0.0609 -		ISM02.4	PPB	9/26/2017 9:54	SMPL
I5332-17	MBDA96	-248	-0.0619 -		ISM02.4	PPB	9/26/2017 9:56	SMPL
I5332-18	MBDA98	1206	0.2958 -		ISM02.4	PPB	9/26/2017 9:58	SMPL
I5332-19	MBDA99	-126	-0.0318 -		ISM02.4	PPB	9/26/2017 10:00	SMPL
CCV006	CCV006	20272	4.9862 -		ISM02.4	PPB	9/26/2017 10:03	SMPL
CCB006	CCB006	-121	-0.0306 -		ISM02.4	PPB	9/26/2017 10:05	SMPL
I5332-20	MBDAA0	86	0.0203 -		ISM02.4	PPB	9/26/2017 10:07	SMPL
I5332-21	MBDAA1	-205	-0.0513 -		ISM02.4	PPB	9/26/2017 10:10	SMPL
I5332-22	MBDA92	-158	-0.0397 -		ISM02.4	PPB	9/26/2017 10:12	SMPL
PB102598BL	PBW003	-84	-0.0215 -		ISM02.4	PPB	9/26/2017 10:14	SMPL
I5355-01	MYADG3	-203	-0.0508 -		ISM02.4	PPB	9/26/2017 10:16	SMPL
I5355-02	MYADG5	-190	-0.0476 -		ISM02.4	PPB	9/26/2017 10:19	SMPL
I5355-03	MYADK2	-161	-0.0405 -		ISM02.4	PPB	9/26/2017 10:21	SMPL
I5355-04	MYADK4	-177	-0.0444 -		ISM02.4	PPB	9/26/2017 10:23	SMPL
I5355-05	MYADG7	-201	-0.0503 -		ISM02.4	PPB	9/26/2017 10:25	SMPL
I5355-06	MYADG7D	-215	-0.0537 -		ISM02.4	PPB	9/26/2017 10:28	SMPL
I5355-07	MYADG7S	3912	0.9615 -		ISM02.4	PPB	9/26/2017 10:30	SMPL
I5355-08	MYADH8	-209	-0.0523 -		ISM02.4	PPB	9/26/2017 10:32	SMPL
I5355-09	MYADM7	-198	-0.0496 -		ISM02.4	PPB	9/26/2017 10:35	SMPL
I5355-10	MYADH1	-140	-0.0353 -		ISM02.4	PPB	9/26/2017 10:37	SMPL

LB90311
INSTRUMENT ID : CV1

I5355-11	MYADH5	-195	-0.0488 -	ISM02.4	PPB	9/26/2017 10:39	SMPL
I5355-12	MYADJ0	-210	-0.0525 -	ISM02.4	PPB	9/26/2017 10:41	SMPL
I5355-13	MYADK9	-198	-0.0496 -	ISM02.4	PPB	9/26/2017 10:44	SMPL
I5355-14	MYADJ6	-156	-0.0392 -	ISM02.4	PPB	9/26/2017 10:46	SMPL
I5355-15	MYADJ8	11549	2.8403 -	ISM02.4	PPB	9/26/2017 10:48	SMPL
I5355-16	MYADM1	-263	-0.0655 -	ISM02.4	PPB	9/26/2017 10:51	SMPL
CCV007	CCV007	20447	5.0292 -	ISM02.4	PPB	9/26/2017 10:53	SMPL
CCB007	CCB007	-83	-0.0213 -	ISM02.4	PPB	9/26/2017 10:55	SMPL
I5355-17	MYADM9	-127	-0.0321 -	ISM02.4	PPB	9/26/2017 10:57	SMPL
PB102599BL	PBW003	0	-0.0008 -	ISM02.4	PPB	9/26/2017 11:00	SMPL
I5356-01	MYADG2	-193	-0.0483 -	ISM02.4	PPB	9/26/2017 11:02	SMPL
I5356-02	MYADG4	-202	-0.0505 -	ISM02.4	PPB	9/26/2017 11:04	SMPL
I5356-03	MYADK1	-201	-0.0503 -	ISM02.4	PPB	9/26/2017 11:06	SMPL
I5356-04	MYADK3	-205	-0.0513 -	ISM02.4	PPB	9/26/2017 11:09	SMPL
I5356-05	MYADG6	-176	-0.0441 -	ISM02.4	PPB	9/26/2017 11:11	SMPL
I5356-06	MYADG6D	-175	-0.0439 -	ISM02.4	PPB	9/26/2017 11:13	SMPL
I5356-07	MYADG6S	3907	0.9603 -	ISM02.4	PPB	9/26/2017 11:16	SMPL
I5356-08	MYADH7	-189	-0.0473 -	ISM02.4	PPB	9/26/2017 11:18	SMPL
I5356-09	MYADM6	-171	-0.0429 -	ISM02.4	PPB	9/26/2017 11:20	SMPL
I5356-10	MYADH0	-150	-0.0377 -	ISM02.4	PPB	9/26/2017 11:22	SMPL
I5356-11	MYADH4	-168	-0.0422 -	ISM02.4	PPB	9/26/2017 11:25	SMPL
I5356-12	MYADH9	-169	-0.0424 -	ISM02.4	PPB	9/26/2017 11:27	SMPL
I5356-13	MYADL0	-163	-0.0409 -	ISM02.4	PPB	9/26/2017 11:29	SMPL
I5356-14	MYADJ5	-195	-0.0488 -	ISM02.4	PPB	9/26/2017 11:32	SMPL
I5356-15	MYADJ7	-129	-0.0326 -	ISM02.4	PPB	9/26/2017 11:34	SMPL
I5356-16	MYADM0	-176	-0.0441 -	ISM02.4	PPB	9/26/2017 11:36	SMPL
I5356-17	MYADM8	-173	-0.0434 -	ISM02.4	PPB	9/26/2017 11:38	SMPL
PB102602BL	PBW003	7	0.0009 -	ISM02.4	PPB	9/26/2017 11:41	SMPL
CCV008	CCV008	20060	4.934 -	ISM02.4	PPB	9/26/2017 11:43	SMPL
CCB008	CCB008	-161	-0.0405 -	ISM02.4	PPB	9/26/2017 11:45	SMPL
I5372-01	MYADF5	-136	-0.0343 -	ISM02.4	PPB	9/26/2017 11:48	SMPL
I5372-02	MYADH3	645	0.1578 -	ISM02.4	PPB	9/26/2017 11:50	SMPL
I5372-03	MYADL3	-226	-0.0564 -	ISM02.4	PPB	9/26/2017 11:52	SMPL
I5372-04	MYADL3D	-232	-0.0579 -	ISM02.4	PPB	9/26/2017 11:54	SMPL
I5372-05	MYADL3S	3910	0.961 -	ISM02.4	PPB	9/26/2017 11:57	SMPL
I5372-06	MYADF9	-260	-0.0648 -	ISM02.4	PPB	9/26/2017 11:59	SMPL
I5372-07	MYADG1	-222	-0.0555 -	ISM02.4	PPB	9/26/2017 12:01	SMPL
I5372-08	MYADL5	-174	-0.0437 -	ISM02.4	PPB	9/26/2017 12:04	SMPL
I5372-09	MYADL7	-230	-0.0574 -	ISM02.4	PPB	9/26/2017 12:06	SMPL
I5372-10	MYADL9	-53	-0.0139 -	ISM02.4	PPB	9/26/2017 12:08	SMPL
PB102603BL	PBW003	20	0.0041 -	ISM02.4	PPB	9/26/2017 12:10	SMPL
I5373-01	MYADF4	-193	-0.0483 -	ISM02.4	PPB	9/26/2017 12:13	SMPL
I5373-02	MYADG0	-226	-0.0564 -	ISM02.4	PPB	9/26/2017 12:15	SMPL
I5373-03	MYADH2	-214	-0.0535 -	ISM02.4	PPB	9/26/2017 12:17	SMPL
I5373-04	MYADL2	-202	-0.0505 -	ISM02.4	PPB	9/26/2017 12:20	SMPL
I5373-05	MYADL2D	-159	-0.04 -	ISM02.4	PPB	9/26/2017 12:22	SMPL
I5373-06	MYADL2S	3833	0.9421 -	ISM02.4	PPB	9/26/2017 12:24	SMPL

LB90311

INSTRUMENT ID : CV1

I5373-07	MYADF8	-260	-0.0648 -	ISM02.4	PPB	9/26/2017 12:26	SMPL
I5373-08	MYADL4	-199	-0.0498 -	ISM02.4	PPB	9/26/2017 12:29	SMPL
I5373-09	MYADL6	-202	-0.0505 -	ISM02.4	PPB	9/26/2017 12:31	SMPL
CCV009	CCV009	20237	4.9776 -	ISM02.4	PPB	9/26/2017 12:33	SMPL
CCB009	CCB009	-144	-0.0363 -	ISM02.4	PPB	9/26/2017 12:36	SMPL
I5373-10	MYADL8	-206	-0.0515 -	ISM02.4	PPB	9/26/2017 12:38	SMPL
CCV010	CCV010	20120	4.9488 -	ISM02.4	PPB	9/26/2017 12:42	SMPL
CCB010	CCB010	-141	-0.0355 -	ISM02.4	PPB	9/26/2017 12:45	SMPL