

245.1


LB90320

INSTRUMENT ID: CY1

Linear

 μ Abs.:

31860



A= 0.0000e+000

B= 3.1148e-004

C= 1.7465e-002

Rho= 0.9999402

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	-107	0.000	-107				
0.2	0.200	0.191	-0.009	558	0.0 %	558				
2.5	2.500	2.479	-0.021	7903	0.0 %	7903				
5.0	5.000	5.060	0.060	16189	0.0 %	16189				
7.5	7.500	7.544	0.044	24164	0.0 %	24164				
10.0	10.000	9.941	-0.059	31860	0.0 %	31860				

LB90320 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Sep 2017 12:57:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-107	-	0.0000	245.1	PPB	26 Sep 2017 12:59:31	S	Std
0.2 - 1	502	558	-	0.2000	245.1	PPB	26 Sep 2017 13:01:39	S	Std
2.5 - 1	52.5	7903	-	2.5000	245.1	PPB	26 Sep 2017 13:03:48	S	Std
5.0 - 1	55	16189	-	5.0000	245.1	PPB	26 Sep 2017 13:05:56	S	Std
7.5 - 1	57.5	24164	-	7.5000	245.1	PPB	26 Sep 2017 13:08:04	S	Std
10.0 - 1	510	31860	-	10.0000	245.1	PPB	26 Sep 2017 13:12:41	S	Std
ICV73 - 1	ICV73	12163	3.8060	-245.1		PPB	26 Sep 2017 13:15:13	U	SMPL
ICB73 - 1	ICB73	-254	-0.0617	-245.1		PPB	26 Sep 2017 13:17:19	U	SMPL
CCV11 - 1	CCV11	15900	4.9700	-245.1		PPB	26 Sep 2017 13:19:26	U	SMPL
CCB11 - 1	CCB11	-248	-0.0598	-245.1		PPB	26 Sep 2017 13:21:32	U	SMPL
CRA - 1	CRA	554	0.1900	-245.1		PPB	26 Sep 2017 13:23:39	U	SMPL
HighStd - 1	HighStd	32343	10.0917	-245.1		PPB	26 Sep 2017 13:25:46	U	SMPL
ChkStd - 1	ChkStd	21658	6.7636	-245.1		PPB	26 Sep 2017 13:27:53	U	SMPL
PB102578BL - 1	PBW	-246	-0.0592	-245.1		PPB	26 Sep 2017 13:30:02	U	SMPL
PB102578BS - 1	LCSW	12713	3.9773	-245.1		PPB	26 Sep 2017 13:32:10	U	SMPL
I5384-01 - 1	092017-TIDO-F-D-M	-247	-0.0595	-245.1		PPB	26 Sep 2017 13:34:18	U	SMPL
I5384-02 - 1	092017-TIDO-F-U-M	-156	-0.0311	-245.1		PPB	26 Sep 2017 13:36:26	U	SMPL
I5384-03 - 1	092017-SIDI-F-D-B	-118	-0.0193	-245.1		PPB	26 Sep 2017 13:38:34	U	SMPL
I5384-04 - 1	092017-SIDI-F-D-S	-125	-0.0215	-245.1		PPB	26 Sep 2017 13:40:43	U	SMPL
I5384-05 - 1	092017-SIDI-F-U-B	-147	-0.0283	-245.1		PPB	26 Sep 2017 13:42:50	U	SMPL
CCV12 - 1	CCV12	15884	4.9651	-245.1		PPB	26 Sep 2017 13:44:57	U	SMPL
CCB12 - 1	CCB12	331	0.1206	-245.1		PPB	26 Sep 2017 13:47:04	U	SMPL
I5384-06 - 1	092017-SIDI-F-U-S	-172	-0.0361	-245.1		PPB	26 Sep 2017 13:49:11	U	SMPL
I5385-01 - 1	092117-SIDI-F-D-B	-130	-0.0230	-245.1		PPB	26 Sep 2017 13:51:18	U	SMPL
I5385-02 - 1	092117-SIDI-F-D-M	-121	-0.0202	-245.1		PPB	26 Sep 2017 13:53:25	U	SMPL
I5385-03 - 1	092117-SIDI-F-D-S	-136	-0.0249	-245.1		PPB	26 Sep 2017 13:55:32	U	SMPL
I5385-04 - 1	092117-SIDI-F-U-B	-78	-0.0068	-245.1		PPB	26 Sep 2017 13:57:40	U	SMPL
I5385-05 - 1	092117-SIDI-F-U-M	-113	-0.0177	-245.1		PPB	26 Sep 2017 13:59:47	U	SMPL
I5385-06 - 1	092117-SIDI-F-U-S	-116	-0.0187	-245.1		PPB	26 Sep 2017 14:01:55	U	SMPL
I5385-07 - 1	092117-TIDO-F-D-M	-126	-0.0218	-245.1		PPB	26 Sep 2017 14:04:03	U	SMPL
I5385-08 - 1	092117-TIDO-F-U-M	-118	-0.0193	-245.1		PPB	26 Sep 2017 14:06:12	U	SMPL
I5385-08DUP - 1	092117-TIDO-F-U-MDUP	-106	-0.0156	-245.1		PPB	26 Sep 2017 14:08:19	U	SMPL
CCV13 - 1	CCV13	15968	4.9912	-245.1		PPB	26 Sep 2017 14:10:26	U	SMPL
CCB13 - 1	CCB13	-228	-0.0536	-245.1		PPB	26 Sep 2017 14:12:33	U	SMPL
I5385-08MS - 1	092117-TIDO-F-U-MMS	15534	4.8560	-245.1		PPB	26 Sep 2017 14:14:40	U	SMPL
I5385-08MSD - 1	092117-TIDO-F-U-MMSD	14986	4.6853	-245.1		PPB	26 Sep 2017 14:16:47	U	SMPL
I5385-08LX5 - 1		-251	-0.0607	-245.1		PPB	26 Sep 2017 14:18:54	U	SMPL
I5385-08A - 1		16071	5.0233	-245.1		PPB	26 Sep 2017 14:21:02	U	SMPL
CCV14 - 1	CCV14	15667	4.8975	-245.1		PPB	26 Sep 2017 14:23:09	U	SMPL
CCB14 - 1	CCB14	-239	-0.0570	-245.1		PPB	26 Sep 2017 14:25:17	U	SMPL