



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB90321

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP58257
Residual chlorine std, 0PPM	WP58252
Residual chlorine std, 0.1PPM	WP58253
Residual chlorine std, 0.2PPM	WP58255
Residual chlorine std, and CCV , 0.4PPM	WP58256
Residual chlorine std, 1.6PPM	WP58258
Residual chlorine ICV-LCS, 0.4PPM	WP58259
Total Chlorine Powder Pillows	W1959

Intercept: -0.0009

Slope: 1.0113

Regression: 0.999968

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	09/26/2017	11:15
2	CAL2	0.1	1	0.000	0.100	0.100	09/26/2017	11:19
3	CAL3	0.2	1	0.000	0.200	0.200	09/26/2017	11:24
4	CAL4	0.4	1	0.000	0.410	0.410	09/26/2017	11:29
5	CAL5	0.8	1	0.000	0.800	0.800	09/26/2017	11:34
6	CAL6	1.6	1	0.000	1.620	1.620	09/26/2017	11:37



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB90321

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.390	0.390	0.387	09/26/2017	11:41
2	ICB				1	0.000	0.000	0.000	0.001	09/26/2017	11:45
3	CCV1			0.4	1	0.000	0.400	0.400	0.396	09/26/2017	11:49
4	CCB1				1	0.000	0.000	0.000	0.001	09/26/2017	11:53
5	LB90321BL	50	50		1	0.000	0.000	0.000	0.001	09/26/2017	11:57
6	LB90321BS	50	50	0.4	1	0.000	0.390	0.390	0.387	09/26/2017	11:59
7	I5274-19	50	50		1	0.000	1.730	1.730	1.712	09/26/2017	12:04
8	CCV2			0.4	1	0.000	0.410	0.410	0.406	09/26/2017	12:10
9	CCB2				1	0.000	0.000	0.000	0.001	09/26/2017	12:15
10	I5274-19				2	0.000	0.860	0.860	0.851	09/26/2017	12:25
11	CCV3			0.4	1	0.000	0.390	0.390	0.387	09/26/2017	12:30
12	CCB3				1	0.000	0.000	0.000	0.001	09/26/2017	12:35