


7470A

LB 90324
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34548



A= 0.0000e+000

B= 2.8765e-004

C= 3.5817e-002

Rho= 0.9999126

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	-29	0.000	-29				
0.2	0.200	0.202	0.002	576	0.0 %	576				
2.5	2.500	2.482	-0.018	8503	0.0 %	8503				
5.0	5.000	4.929	-0.071	17011	0.0 %	17011				
7.5	7.500	7.586	0.086	26249	0.0 %	26249				
10.0	10.000	9.974	-0.026	34548	0.0 %	34548				

LB90324 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Sep 2017 14:40:10					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7470A	PPB	26 Sep 2017 14:41:52	S	Std
0.2 - 1	50.2	576	-	0.2000	7470A	PPB	26 Sep 2017 14:44:00	S	Std
2.5 - 1	52.5	8503	-	2.5000	7470A	PPB	26 Sep 2017 14:46:08	S	Std
5.0 - 1	55	17011	-	5.0000	7470A	PPB	26 Sep 2017 14:48:17	S	Std
7.5 - 1	57.5	26249	-	7.5000	7470A	PPB	26 Sep 2017 14:50:25	S	Std
10.0 - 1	510	34548	-	10.0000	7470A	PPB	26 Sep 2017 14:52:32	S	Std
ICV74 - 1	ICV74	13781	4.0000	-7470A	PPB	26 Sep 2017 14:55:12	U	SMPL	
ICB74 - 1	ICB74	-177	-0.0151	-7470A	PPB	26 Sep 2017 14:57:19	U	SMPL	
CCV15 - 1	CCV15	17383	5.0361	-7470A	PPB	26 Sep 2017 14:59:25	U	SMPL	
CCB15 - 1	CCB15	-175	-0.0145	-7470A	PPB	26 Sep 2017 15:01:32	U	SMPL	
CRA - 1	CRA	523	0.1863	-7470A	PPB	26 Sep 2017 15:03:39	U	SMPL	
HighStd - 1	HighStd	35257	10.1777	-7470A	PPB	26 Sep 2017 15:05:46	U	SMPL	
ChkStd - 1	ChkStd	24124	6.9752	-7470A	PPB	26 Sep 2017 15:07:53	U	SMPL	
PB102579BL - 1	PBW	-246	-0.0349	-7470A	PPB	26 Sep 2017 15:10:01	U	SMPL	
PB102579BS - 1	LCSW	13375	3.8832	-7470A	PPB	26 Sep 2017 15:12:10	U	SMPL	
PB102579TB - 1	PB102550TB	-149	-0.0070	-7470A	PPB	26 Sep 2017 15:14:17	U	SMPL	
I5380-03 - 1	SU-04-092217	-54	0.0203	-7470A	PPB	26 Sep 2017 15:16:26	U	SMPL	
I5394-03 - 1	15-50	-39	0.0246	-7470A	PPB	26 Sep 2017 15:18:34	U	SMPL	
I5394-06 - 1	20-41	-26	0.0283	-7470A	PPB	26 Sep 2017 15:20:42	U	SMPL	
I5394-09 - 1	RT-3992	-52	0.0209	-7470A	PPB	26 Sep 2017 15:22:49	U	SMPL	
CCV16 - 1	CCV16	17088	4.9513	-7470A	PPB	26 Sep 2017 15:24:56	U	SMPL	
CCB16 - 1	CCB16	-149	-0.0070	-7470A	PPB	26 Sep 2017 15:27:03	U	SMPL	
I5395-03 - 1	TP-5-(0-5)	-56	0.0197	-7470A	PPB	26 Sep 2017 15:29:10	U	SMPL	
I5395-06 - 1	TP-4-(0-5)	-42	0.0237	-7470A	PPB	26 Sep 2017 15:31:17	U	SMPL	
I5395-09 - 1	TP-3-(0-5)	2	0.0364	-7470A	PPB	26 Sep 2017 15:33:24	U	SMPL	
I5395-12 - 1	TP-2-(0-5)	-34	0.0260	-7470A	PPB	26 Sep 2017 15:35:31	U	SMPL	
I5395-15 - 1	TP-1-(0-5)	-19	0.0304	-7470A	PPB	26 Sep 2017 15:37:39	U	SMPL	
I5395-18 - 1	TP-6-(0-5)	-23	0.0292	-7470A	PPB	26 Sep 2017 15:39:46	U	SMPL	
I5395-18DUP - 1	TP-6-(0-5)DUP	-13	0.0321	-7470A	PPB	26 Sep 2017 15:41:55	U	SMPL	
I5395-18MS - 1	TP-6-(0-5)MS	13203	3.8337	-7470A	PPB	26 Sep 2017 15:44:03	U	SMPL	
I5395-18MSD - 1	TP-6-(0-5)MSD	13107	3.8061	-7470A	PPB	26 Sep 2017 15:46:12	U	SMPL	
PB102580BL - 1	PBW	-115	0.0027	-7470A	PPB	26 Sep 2017 15:48:18	U	SMPL	
CCV17 - 1	CCV17	17222	4.9898	-7470A	PPB	26 Sep 2017 15:50:25	U	SMPL	
CCB17 - 1	CCB17	-171	-0.0134	-7470A	PPB	26 Sep 2017 15:52:33	U	SMPL	
PB102580BS - 1	LCSW	13746	3.9899	-7470A	PPB	26 Sep 2017 15:57:14	U	SMPL	
I5381-01 - 1	20170921-092006	-166	-0.0119	-7470A	PPB	26 Sep 2017 15:59:21	U	SMPL	
I5381-02 - 1	20170921-092529	-36	0.0255	-7470A	PPB	26 Sep 2017 16:01:27	U	SMPL	
I5381-03 - 1	20170921-092732	-73	0.0148	-7470A	PPB	26 Sep 2017 16:03:34	U	SMPL	
I5381-04 - 1	20170921-102649	-55	0.0200	-7470A	PPB	26 Sep 2017 16:05:42	U	SMPL	
I5381-05 - 1	20170921-103110	-125	-0.0001	-7470A	PPB	26 Sep 2017 16:07:49	U	SMPL	
I5381-06 - 1	20170921-103319	-114	0.0030	-7470A	PPB	26 Sep 2017 16:09:57	U	SMPL	
I5381-07 - 1	20170921-135234	-108	0.0048	-7470A	PPB	26 Sep 2017 16:12:06	U	SMPL	
I5381-08 - 1	20170921-135654	-83	0.0119	-7470A	PPB	26 Sep 2017 16:14:14	U	SMPL	
I5381-09 - 1	20170921-135850	-117	0.0022	-7470A	PPB	26 Sep 2017 16:16:22	U	SMPL	
CCV18 - 1	CCV18	17224	4.9904	-7470A	PPB	26 Sep 2017 16:18:29	U	SMPL	
CCB18 - 1	CCB18	-213	-0.0255	-7470A	PPB	26 Sep 2017 16:20:36	U	SMPL	
I5381-10 - 1	20170921-150548	-95	0.0085	-7470A	PPB	26 Sep 2017 16:22:44	U	SMPL	
I5381-11 - 1	20170921-151059	-95	0.0085	-7470A	PPB	26 Sep 2017 16:24:51	U	SMPL	
I5381-12 - 1	20170921-151227	-122	0.0007	-7470A	PPB	26 Sep 2017 16:26:58	U	SMPL	
I5397-01 - 1	20170922-091918	-61	0.0183	-7470A	PPB	26 Sep 2017 16:29:05	U	SMPL	
I5397-02 - 1	20170922-092451	-91	0.0096	-7470A	PPB	26 Sep 2017 16:31:13	U	SMPL	
I5397-03 - 1	20170922-092946	-125	-0.0001	-7470A	PPB	26 Sep 2017 16:33:21	U	SMPL	
I5397-04 - 1	20170922-120119	-44	0.0232	-7470A	PPB	26 Sep 2017 16:35:29	U	SMPL	
I5397-05 - 1	20170922-120416	-110	0.0042	-7470A	PPB	26 Sep 2017 16:37:37	U	SMPL	
I5397-06 - 1	20170922-120657	-113	0.0033	-7470A	PPB	26 Sep 2017 16:39:46	U	SMPL	
I5397-06DUP - 1	20170922-120657DUP	-119	0.0016	-7470A	PPB	26 Sep 2017 16:41:53	U	SMPL	
CCV19 - 1	CCV19	16973	4.9182	-7470A	PPB	26 Sep 2017 16:44:01	U	SMPL	
CCB19 - 1	CCB19	-211	-0.0249	-7470A	PPB	26 Sep 2017 16:46:08	U	SMPL	
I5397-06MS - 1	20170922-120657MS	14528	4.2149	-7470A	PPB	26 Sep 2017 16:48:15	U	SMPL	
I5397-06MSD - 1	20170922-120657MSD	14775	4.2859	-7470A	PPB	26 Sep 2017 16:50:23	U	SMPL	
PB102581BL - 1	PBW	-143	-0.0053	-7470A	PPB	26 Sep 2017 16:52:30	U	SMPL	
PB102581BS - 1	LCSW	13208	3.8352	-7470A	PPB	26 Sep 2017 16:54:37	U	SMPL	
PB102581TB - 1	PB102545TB	-261	-0.0393	-7470A	PPB	26 Sep 2017 16:56:45	U	SMPL	
I5370-08 - 1	G-34-1.5-3	-82	0.0122	-7470A	PPB	26 Sep 2017 16:58:53	U	SMPL	
I5370-08DUP - 1	G-34-1.5-3DUP	-71	0.0154	-7470A	PPB	26 Sep 2017 17:01:01	U	SMPL	
I5370-08MS - 1	G-34-1.5-3MS	13585	3.9436	-7470A	PPB	26 Sep 2017 17:03:09	U	SMPL	
I5370-08MSD - 1	G-34-1.5-3MSD	13750	3.9911	-7470A	PPB	26 Sep 2017 17:05:18	U	SMPL	
CCV20 - 1	CCV20	17152	4.9697	-7470A	PPB	26 Sep 2017 17:07:25	U	SMPL	
CCB20 - 1	CCB20	-220	-0.0275	-7470A	PPB	26 Sep 2017 17:09:33	U	SMPL	
I5395-18LX5 - 1		-102	0.0065	-7470A	PPB	26 Sep 2017 17:11:41	U	SMPL	
I5395-18A - 1		14145	4.1047	-7470A	PPB	26 Sep 2017 17:13:48	U	SMPL	
I5397-06LX5 - 1		-206	-0.0234	-7470A	PPB	26 Sep 2017 17:15:56	U	SMPL	
I5397-06A - 1		15389	4.4625	-7470A	PPB	26 Sep 2017 17:18:03	U	SMPL	
I5370-08LX5 - 1		-203	-0.0226	-7470A	PPB	26 Sep 2017 17:20:11	U	SMPL	
I5370-08A - 1		13739	3.9879	-7470A	PPB	26 Sep 2017 17:22:18	U	SMPL	
CCV21 - 1	CCV21	17120	4.9605	-7470A	PPB	26 Sep 2017 17:24:26	U	SMPL	
CCB21 - 1	CCB21	-169	-0.0128	-7470A	PPB	26 Sep 2017 17:26:34	U	SMPL	