

Analysis Method: 9034

ANALYST: ketankumar

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB90328

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2206
IODINE SOLUTION .025N 1L	W2251
Starch Solution, 4L	W2121

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB102620BL		1	50	50	2.00	0.00	1.70	1.70	0.30	0.00	0.00	09/26/2017	12:52
2	I5429-02		1	50	50	2.00	1.70	3.38	1.68	0.32	0.02	0.16	09/26/2017	12:55
3	I5429-02DUP		1	50	50	2.00	3.38	5.04	1.66	0.34	0.04	0.32	09/26/2017	12:58

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 \* Normality1) - ((T1 - Value Corrected With Blank) \* Normality2)) \* Constant / Initial Volume