


7471A

LB90334
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

41206



A= 0.0000e+000

B= 2.4166e-004

C= 5.5166e-003

Rho= 0.9999546

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-115	0.000	-115				
0.2	0.200	0.193	-0.007	774	0.0 %	774				
2.5	2.500	2.542	0.042	10497	0.0 %	10497				
5.0	5.000	4.970	-0.030	20545	0.0 %	20545				
7.5	7.500	7.554	0.054	31234	0.0 %	31234	0			
10.0	10.000	9.963	-0.037	41206	0.0 %	41206				

LB90334 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 26 Sep 2017 17:56:02					
Sample ID	Extended ID	µ Abs	Conc.	Std CondMethod	Units	Date	Type	Type	
0.0 - 1	50	-115	-	0.00007471A	PPB	26 Sep 2017 18:02:27	S	Std	
0.2 - 1	50.2	774	-	0.20007471A	PPB	26 Sep 2017 18:04:45	S	Std	
2.5 - 1	52.5	10497	-	2.50007471A	PPB	26 Sep 2017 18:07:03	S	Std	
5.0 - 1	55	20545	-	5.00007471A	PPB	26 Sep 2017 18:09:20	S	Std	
7.5 - 1	57.5	31234	-	7.50007471A	PPB	26 Sep 2017 18:11:37	S	Std	
10.0 - 1	510	41206	-	10.00007471A	PPB	26 Sep 2017 18:19:07	S	Std	
ICV75 - 1	ICV75	15288	3.7000	-7471A	PPB	26 Sep 2017 18:21:48	U	SMPL	
ICB75 - 1	ICB75	-273	-0.0605	-7471A	PPB	26 Sep 2017 18:24:04	U	SMPL	
CCV22 - 1	CCV22	20822	5.0374	-7471A	PPB	26 Sep 2017 18:26:20	U	SMPL	
CCB22 - 1	CCB22	-277	-0.0614	-7471A	PPB	26 Sep 2017 18:28:36	U	SMPL	
CRA - 1	CRA	783	0.1947	-7471A	PPB	26 Sep 2017 18:30:52	U	SMPL	
HighStd - 1	HighStd	40432	9.7764	-7471A	PPB	26 Sep 2017 18:33:08	U	SMPL	
ChkStd - 1	ChkStd	27771	6.7167	-7471A	PPB	26 Sep 2017 18:35:24	U	SMPL	
PB102604BL - 1	PBS	-239	-0.0522	-7471A	PPB	26 Sep 2017 18:40:16	U	SMPL	
PB102604BS - 1	LCSS	13385	3.2402	-7471A	PPB	26 Sep 2017 18:42:33	U	SMPL	
IS380-01 - 1	SU-04-092217	1969	0.4813	-7471A	PPB	26 Sep 2017 18:44:50	U	SMPL	
IS386-01 - 1	SASP2P-ENV-116-2(0-5)	1368	0.3361	-7471A	PPB	26 Sep 2017 18:47:07	U	SMPL	
IS386-02 - 1	SASP2P-ENV-116-2(5-10)	2296	0.5604	-7471A	PPB	26 Sep 2017 18:49:24	U	SMPL	
IS388-01 - 1	FK-PS-01-016	406	0.1036	-7471A	PPB	26 Sep 2017 18:51:42	U	SMPL	
IS391-01 - 1	FK-PS-01-015	325	0.0841	-7471A	PPB	26 Sep 2017 18:53:58	U	SMPL	
CCV23 - 1	CCV23	21177	5.1232	-7471A	PPB	26 Sep 2017 18:56:14	U	SMPL	
CCB23 - 1	CCB23	-209	-0.0450	-7471A	PPB	26 Sep 2017 18:58:29	U	SMPL	
IS391-01DUP - 1	FK-PS-01-015DUP	325	0.0841	-7471A	PPB	26 Sep 2017 19:03:28	U	SMPL	
IS391-01MS - 1	FK-PS-01-015MS	15179	3.6737	-7471A	PPB	26 Sep 2017 19:05:44	U	SMPL	
IS391-01MSD - 1	FK-PS-01-015MSD	14881	3.6017	-7471A	PPB	26 Sep 2017 19:08:00	U	SMPL	
IS394-01 - 1	15-50	2669	0.6505	-7471A	PPB	26 Sep 2017 19:10:16	U	SMPL	
IS394-04 - 1	20-41	2503	0.6104	-7471A	PPB	26 Sep 2017 19:12:34	U	SMPL	
IS394-07 - 1	RT-3992	1483	0.3639	-7471A	PPB	26 Sep 2017 19:14:51	U	SMPL	
IS395-01 - 1	TP-5(0-5)	2315	0.5650	-7471A	PPB	26 Sep 2017 19:17:08	U	SMPL	
IS395-04 - 1	TP-4(0-5)	2387	0.5824	-7471A	PPB	26 Sep 2017 19:19:25	U	SMPL	
IS395-07 - 1	TP-3(0-5)	2273	0.5548	-7471A	PPB	26 Sep 2017 19:21:44	U	SMPL	
IS395-10 - 1	TP-2(0-5)	2031	0.4963	-7471A	PPB	26 Sep 2017 19:23:59	U	SMPL	
CCV24 - 1	CCV24	21334	5.1611	-7471A	PPB	26 Sep 2017 19:26:16	U	SMPL	
CCB24 - 1	CCB24	-224	-0.0486	-7471A	PPB	26 Sep 2017 19:28:32	U	SMPL	
IS395-13 - 1	TP-1(0-5)	1970	0.4816	-7471A	PPB	26 Sep 2017 19:30:49	U	SMPL	
IS395-16 - 1	TP-6(0-5)	1794	0.4391	-7471A	PPB	26 Sep 2017 19:33:06	U	SMPL	
IS396-01 - 1	PAMT-COMP-A	893	0.2213	-7471A	PPB	26 Sep 2017 19:35:22	U	SMPL	
IS396-02 - 1	PAMT-COMP-B	4067	0.9884	-7471A	PPB	26 Sep 2017 19:37:38	U	SMPL	
IS396-03 - 1	PAMT-COMP-E	14566	3.5256	-7471A	PPB	26 Sep 2017 19:39:54	U	SMPL	
IS396-04 - 1	PAMT-COMP-A-DUP	970	0.2399	-7471A	PPB	26 Sep 2017 19:42:11	U	SMPL	
PB102605BL - 1	PBS	-73	-0.0121	-7471A	PPB	26 Sep 2017 19:44:28	U	SMPL	
PB102605BS - 1	LCSS	13901	3.3649	-7471A	PPB	26 Sep 2017 19:46:46	U	SMPL	
IS418-01 - 1	SASP2P-ENV-119-2(0-5)	245	0.0647	-7471A	PPB	26 Sep 2017 19:49:04	U	SMPL	
IS418-01DUP - 1	SASP2P-ENV-119-2(0-5)DUP	326	0.0843	-7471A	PPB	26 Sep 2017 19:53:54	U	SMPL	
CCV25 - 1	CCV25	21599	5.2252	-7471A	PPB	26 Sep 2017 19:56:09	U	SMPL	
CCB25 - 1	CCB25	-228	-0.0496	-7471A	PPB	26 Sep 2017 19:58:26	U	SMPL	
IS418-01MS - 1	SASP2P-ENV-119-2(0-5)MS	15474	3.7450	-7471A	PPB	26 Sep 2017 20:00:43	U	SMPL	
IS418-01MSD - 1	SASP2P-ENV-119-2(0-5)MSD	15205	3.6800	-7471A	PPB	26 Sep 2017 20:02:59	U	SMPL	
IS418-02 - 1	SASP2P-ENV-119-2(5-10)	498	0.1259	-7471A	PPB	26 Sep 2017 20:05:16	U	SMPL	
IS418-03 - 1	SASP2P-ENV-125-17(20-30)	32	0.0132	-7471A	PPB	26 Sep 2017 20:07:32	U	SMPL	
IS419-01 - 1	SASP2P-TP-106V(0-5)	3618	0.8798	-7471A	PPB	26 Sep 2017 20:09:49	U	SMPL	
IS421-01 - 1	JC-DRUM	81537	19.7099	-7471A	PPB	26 Sep 2017 20:12:06	U	SMPL	
IS423-01 - 1	SA-06-092517-A	965	0.2387	-7471A	PPB	26 Sep 2017 20:14:23	U	SMPL	
IS423-04 - 1	SA-06-092517-B	1464	0.3593	-7471A	PPB	26 Sep 2017 20:16:43	U	SMPL	
IS428-01 - 1	A-B-C-DRUM-COMP	6413	1.5553	-7471A	PPB	26 Sep 2017 20:19:01	U	SMPL	
PB102612BL - 1	PBS	-85	-0.0150	-7471A	PPB	26 Sep 2017 20:21:17	U	SMPL	
CCV26 - 1	CCV26	21050	5.0925	-7471A	PPB	26 Sep 2017 20:23:34	U	SMPL	
CCB26 - 1	CCB26	-231	-0.0503	-7471A	PPB	26 Sep 2017 20:25:50	U	SMPL	
PB102612BS - 1	LCSS	14371	3.4784	-7471A	PPB	26 Sep 2017 20:28:07	U	SMPL	
IS360-01 - 1	SP-1-A-COMP	22814	5.5188	-7471A	PPB	26 Sep 2017 20:30:24	U	SMPL	
IS360-03 - 1	SP-1-B-COMP	43589	10.5393	-7471A	PPB	26 Sep 2017 20:32:40	U	SMPL	
IS360-05 - 1	SP-1-C-COMP	23234	5.6203	-7471A	PPB	26 Sep 2017 20:34:58	U	SMPL	
IS360-07 - 1	SP-1-D-COMP	24154	5.8426	-7471A	PPB	26 Sep 2017 20:37:16	U	SMPL	
IS360-09 - 1	SP-1-E-COMP	24719	5.9792	-7471A	PPB	26 Sep 2017 20:39:36	U	SMPL	
IS360-11 - 1	SP-1-F-COMP	26773	6.4755	-7471A	PPB	26 Sep 2017 20:41:54	U	SMPL	
IS360-13 - 1	SP-1-G-COMP	35248	8.5236	-7471A	PPB	26 Sep 2017 20:44:12	U	SMPL	
IS360-15 - 1	SP-1-H-COMP	25822	6.2457	-7471A	PPB	26 Sep 2017 20:46:30	U	SMPL	
IS362-01 - 1	SP-2-A-COMP	14157	3.4267	-7471A	PPB	26 Sep 2017 20:48:48	U	SMPL	
CCV27 - 1	CCV27	20832	5.0398	-7471A	PPB	26 Sep 2017 20:51:06	U	SMPL	
CCB27 - 1	CCB27	-289	-0.0643	-7471A	PPB	26 Sep 2017 20:53:22	U	SMPL	
IS362-03 - 1	SP-2-B-COMP	14580	3.5289	-7471A	PPB	26 Sep 2017 20:55:39	U	SMPL	
IS362-05 - 1	SP-2-C-COMP	18328	4.4347	-7471A	PPB	26 Sep 2017 20:57:56	U	SMPL	
IS363-01 - 1	SP-3-A-COMP	52726	12.7474	-7471A	PPB	26 Sep 2017 21:00:13	U	SMPL	
IS363-03 - 1	SP-3-B-COMP	41444	10.0209	-7471A	PPB	26 Sep 2017 21:02:31	U	SMPL	
IS363-05 - 1	SP-3-C-COMP	47384	11.4564	-7471A	PPB	26 Sep 2017 21:04:50	U	SMPL	
IS363-07 - 1	SP-3-D-COMP	43094	10.4197	-7471A	PPB	26 Sep 2017 21:07:08	U	SMPL	
IS363-09 - 1	SP-3-E-COMP	41746	10.0939	-7471A	PPB	26 Sep 2017 21:09:28	U	SMPL	
IS363-11 - 1	SP-3-F-COMP	40699	9.8409	-7471A	PPB	26 Sep 2017 21:11:47	U	SMPL	
IS363-13 - 1	SP-3-G-COMP	47096	11.3868	-7471A	PPB	26 Sep 2017 21:14:07	U	SMPL	
IS363-15 - 1	SP-3-H-COMP	48747	11.7858	-7471A	PPB	26 Sep 2017 21:16:26	U	SMPL	
CCV28 - 1	CCV28	20398	4.9349	-7471A	PPB	26 Sep 2017 21:18:45	U	SMPL	
CCB28 - 1	CCB28	-272	-0.0602	-7471A	PPB	26 Sep 2017 21:21:05	U	SMPL	
IS363-15DUP - 1	SP-3-H-COMPDUP	47931	11.5886	-7471A	PPB	26 Sep 2017 21:23:22	U	SMPL	
IS363-15MSD - 1	SP-3-H-COMPMSD	62736	15.1664	-7471A	PPB	26 Sep 2017 21:27:58	U	SMPL	
IS363-15MS - 1	SP-3-H-COMPMS	62490	15.1070	-7471A	PPB	26 Sep 2017 21:30:40	U	SMPL	
IS391-01LX5 - 1		-315	-0.0706	-7471A	PPB	26 Sep 2017 21:33:26	U	SMPL	
IS391-01A - 1		17203	4.1628	-7471A	PPB	26 Sep 2017 21:35:44	U	SMPL	
IS418-01LX5 - 1		-118	-0.0230	-7471A	PPB	26 Sep 2017 21:40:50	U	SMPL	
IS418-01A - 1		17046	4.1249	-7471A	PPB	26 Sep 2017 21:43:10	U	SMPL	
IS363-15LX5 - 1		8913	2.1594	-7471A	PPB	26 Sep 2017 21:45:28	U	SMPL	
IS363-15A - 1		64137	15.5050	-7471A	PPB	26 Sep 2017 21:47:45	U	SMPL	
CCV29 - 1	CCV29	19482	4.7136	-7471A	PPB	26 Sep 2017 21:50:02	U	SMPL	
CCB29 - 1	CCB29	-346	-0.0781	-7471A	PPB	26 Sep 2017 21:52:21	U	SMPL	

LB90334 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 26 Sep 2017 17:56:02					
Sample ID	Extended ID	µ Abs	Conc.	Strnd Conc	Method	Units	Date	Type	Type
I5421-01DLX5 - 1		16245	3.9313		-7471A	PPB	26 Sep 2017 21:58:57	U	SMPL
I5360-03DLX2 - 1		22975	5.5577		-7471A	PPB	26 Sep 2017 22:01:14	U	SMPL
I5363-01DLX2 - 1		27618	6.6797		-7471A	PPB	26 Sep 2017 22:03:32	U	SMPL
I5363-03DLX2 - 1		22065	5.3378		-7471A	PPB	26 Sep 2017 22:05:50	U	SMPL
I5363-05DLX2 - 1		25136	6.0799		-7471A	PPB	26 Sep 2017 22:08:09	U	SMPL
I5363-07DLX2 - 1		23258	5.6261		-7471A	PPB	26 Sep 2017 22:10:26	U	SMPL
I5363-09DLX2 - 1		22537	5.4518		-7471A	PPB	26 Sep 2017 22:12:44	U	SMPL
I5363-13DLX2 - 1		24019	5.8100		-7471A	PPB	26 Sep 2017 22:15:02	U	SMPL
CCV30 - 1	CCV30	20868	5.0485		-7471A	PPB	26 Sep 2017 22:17:20	U	SMPL
CCB30 - 1	CCB30	-290	-0.0646		-7471A	PPB	26 Sep 2017 22:19:39	U	SMPL
I5363-15DLX2 - 1		24920	6.0277		-7471A	PPB	26 Sep 2017 22:21:57	U	SMPL
I5363-15DUPDLX2 - 1		23686	5.7295		-7471A	PPB	26 Sep 2017 22:24:14	U	SMPL
I5363-15MSDLX2 - 1		31250	7.5574		-7471A	PPB	26 Sep 2017 22:26:31	U	SMPL
I5363-15MSDDLX2 - 1		30073	7.2730		-7471A	PPB	26 Sep 2017 22:28:48	U	SMPL
I5363-15LDLX10 - 1		5138	1.2472		-7471A	PPB	26 Sep 2017 22:31:06	U	SMPL
I5363-15ADLX2 - 1		32323	7.8168		-7471A	PPB	26 Sep 2017 22:33:25	U	SMPL
CCV31 - 1	CCV31	19689	4.7636		-7471A	PPB	26 Sep 2017 22:35:42	U	SMPL
CCB31 - 1	CCB31	-241	-0.0527		-7471A	PPB	26 Sep 2017 22:38:01	U	SMPL